



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB103102

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	05/30/2019	12:25
2	K3085-02		1	8	100.00	>212.0	>212.0	05/30/2019	12:50
3	K3085-02DUP		1	8	100.00	>212.0	>212.0	05/30/2019	13:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 103102

WorkList Name : FP53019

WorkList ID : 126741

Date : 05-30-2019 09:32:49

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/11/2019	Water	K3085-02	Flash Point	Cool 4 deg C	DVIR01	TANK-3		05/29/2019	1010A

Date/Time 05/30/19 12:00
 Received by: KS
 Relinquished by: CP

Date/Time 05/30/19 13:25
 Received by: CP
 Relinquished by: KS