



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB103106

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH MER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	05/30/2019	18:00
2	CAL2	1	Water	NA	NA	20.1	6.99	05/30/2019	18:01
3	CAL3	1	Water	NA	NA	20.1	10.02	05/30/2019	18:02
4	ICV	1	Water	NA	NA	20.1	7.01	05/30/2019	18:03
5	CCV1	1	Water	NA	NA	20.1	2.01	05/30/2019	18:05
6	K2639-06	1	Solid	20.02	20	23.0	9.03	05/30/2019	18:06
7	K2639-06DUP	1	Solid	20.03	20	23.1	9.04	05/30/2019	18:07
8	K2639-08	1	Solid	20.02	20	23.0	8.19	05/30/2019	18:09
9	K3104-11	1	Solid	20.01	20	22.7	6.70	05/30/2019	18:10
10	K3105-02	1	Solid	20.03	20	21.6	7.59	05/30/2019	18:11
11	K3105-04	1	Solid	20.04	20	21.5	7.33	05/30/2019	18:12
12	K3105-06	1	Solid	20.03	20	21.7	8.36	05/30/2019	18:14
13	K3105-08	1	Solid	20.05	20	21.8	6.95	05/30/2019	18:15
14	K3105-10	1	Solid	20.02	20	21.9	5.90	05/30/2019	18:16
15	K3124-06	1	Solid	20.03	20	22.4	7.75	05/30/2019	18:18
16	CCV2	1	Water	NA	NA	20.1	12.02	05/30/2019	18:20
17	K3124-10	1	Solid	20.04	20	22.0	6.35	05/30/2019	18:21
18	CCV3	1	Water	NA	NA	20.1	2.02	05/30/2019	18:24

18103106

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity k3105

WorkList ID : 126773

Date : 05-30-2019 15:12:00

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/30/2019	Solid	K2639-06	Corrosivity	Cool 4 deg C	PSEG05		NB-07-053019-A	05/30/2019	9045D
05/30/2019	Solid	K2639-08	Corrosivity	Cool 4 deg C	PSEG05		NB-07-053019-B	05/30/2019	9045D
05/30/2019	Solid	K3104-11	Corrosivity	Cool 4 deg C	PSEG01	O31	RBR-251-688	05/30/2019	9045D
05/30/2019	Solid	K3105-02	Corrosivity	Cool 4 deg C	TULL02	R31	101-102	05/30/2019	9045D
05/30/2019	Solid	K3105-04	Corrosivity	Cool 4 deg C	TULL02	R31	APPLE-TRACK	05/30/2019	9045D
05/30/2019	Solid	K3105-06	Corrosivity	Cool 4 deg C	TULL02	R31	TRIANGLE	05/30/2019	9045D
05/30/2019	Solid	K3105-08	Corrosivity	Cool 4 deg C	TULL02	R31	8-9-TRACK	05/30/2019	9045D
05/30/2019	Solid	K3105-10	Corrosivity	Cool 4 deg C	TULL02	R31	66-67-TRACK	05/30/2019	9045D
05/30/2019	Solid	K3124-06	Corrosivity	Cool 4 deg C	PSEG01	O31	TP-25-S	05/30/2019	9045D
05/30/2019	Solid	K3124-10	Corrosivity	Cool 4 deg C	PSEG01	O31	TP-25-D	05/30/2019	9045D

Date/Time 05.30.19 16:15
Received by: Jg
Relinquished by: PC

Date/Time 05.30.19 18:35
Received by: AC
Relinquished by: Jg