



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB103107

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	05/30/2019	18:00
2	CAL2	1	Water	NA	NA	20.1	6.99	05/30/2019	18:01
3	CAL3	1	Water	NA	NA	20.1	10.02	05/30/2019	18:02
4	ICV	1	Water	NA	NA	20.1	7.01	05/30/2019	18:03
5	CCV1	1	Water	NA	NA	20.1	2.01	05/30/2019	18:25
6	K3104-02	1	Water	NA	NA	14.2	6.40	05/30/2019	18:26
7	K3104-02DUP	1	Water	NA	NA	14.3	6.41	05/30/2019	18:27
8	K3104-05	1	Water	NA	NA	13.0	6.60	05/30/2019	18:30
9	K3104-08	1	Water	NA	NA	13.2	6.40	05/30/2019	18:31
10	CCV2	1	Water	NA	NA	20.2	12.01	05/30/2019	18:32

19103102X

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity k3141

WorkList ID : 126774

Date : 05-30-2019 15:13:52

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/30/2019	Water	K3104-02	Corrosivity	Cool 4 deg C	PSEG01	O31	251810	05/30/2019	9040C
05/30/2019	Water	K3104-05	Corrosivity	Cool 4 deg C	PSEG01	O31	SUMP	05/30/2019	9040C
05/30/2019	Water	K3104-08	Corrosivity	Cool 4 deg C	PSEG01	O31	PIT-WATER	05/30/2019	9040C

Date/Time 05.30.19 16:15
Received by: Jy
Relinquished by: PC

Date/Time 05.30.19 18:35
Received by: PC
Relinquished by: Jy