



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB103108

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	05/30/2019	18:00
2	CAL2	1	Water	NA	NA	20.1	6.99	05/30/2019	18:01
3	CAL3	1	Water	NA	NA	20.1	10.02	05/30/2019	18:02
4	ICV	1	Water	NA	NA	20.1	7.01	05/30/2019	18:03
5	CCV1	1	Water	NA	NA	20.1	2.01	05/30/2019	18:25
6	K3093-01	1	Water	NA	NA	20.2	6.03	05/30/2019	18:26
7	K3093-01DUP	1	Water	NA	NA	20.1	6.04	05/30/2019	18:27
8	K3093-02	1	Water	NA	NA	20.4	6.41	05/30/2019	18:30
9	K3104-01	1	Water	NA	NA	14.2	6.40	05/30/2019	18:31
10	K3104-04	1	Water	NA	NA	13.0	6.60	05/30/2019	18:32
11	K3104-07	1	Water	NA	NA	13.2	6.40	05/30/2019	18:34
12	K3125-01	1	Water	NA	NA	19.6	6.48	05/30/2019	18:35
13	K3125-02	1	Water	NA	NA	19.7	5.71	05/30/2019	18:36
14	K3125-03	1	Water	NA	NA	19.7	6.23	05/30/2019	18:27
15	K3125-04	1	Water	NA	NA	19.9	5.98	05/30/2019	18:38
16	CCV2	1	Water	NA	NA	20.1	12.02	05/30/2019	18:39
17	K3125-05	1	Water	NA	NA	19.4	6.34	05/30/2019	18:41
18	K3125-06	1	Water	NA	NA	19.8	6.29	05/30/2019	18:42
19	K3125-07	1	Water	NA	NA	20.2	6.15	05/30/2019	18:43
20	K3125-08	1	Water	NA	NA	20.0	6.32	05/30/2019	18:45

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	K3125-09	1	Water	NA	NA	20.2	5.52	05/30/2019	18:46
22	K3125-10	1	Water	NA	NA	20.1	6.86	05/30/2019	18:47
23	K3125-11	1	Water	NA	NA	19.7	6.42	05/30/2019	18:48
24	K3125-12	1	Water	NA	NA	20.3	5.85	05/30/2019	18:49
25	K3125-13	1	Water	NA	NA	20.5	5.75	05/30/2019	18:50
26	CCV3	1	Water	NA	NA	20.1	2.02	05/30/2019	18:51

2B 103108

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K3141

WorkList ID : 126775

Date : 05-30-2019 15:14:30

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/30/2019	Water	K3093-01	pH	Cool 4 deg C	PSEG01	K11	13907	05/30/2019	9040C
05/30/2019	Water	K3093-02	pH	Cool 4 deg C	PSEG01	K11	13908	05/30/2019	9040C
05/30/2019	Water	K3104-01	pH	Cool 4 deg C	PSEG01	O31	251810	05/30/2019	9040C
05/30/2019	Water	K3104-04	pH	Cool 4 deg C	PSEG01	O31	SUMP	05/30/2019	9040C
05/30/2019	Water	K3104-07	pH	Cool 4 deg C	PSEG01	O31	PIT-WATER	05/30/2019	9040C
05/30/2019	Water	K3125-01	pH	Cool 4 deg C	PSEG01	O31	13887	05/30/2019	9040C
05/30/2019	Water	K3125-02	pH	Cool 4 deg C	PSEG01	O31	13888	05/30/2019	9040C
05/30/2019	Water	K3125-03	pH	Cool 4 deg C	PSEG01	O31	13891	05/30/2019	9040C
05/30/2019	Water	K3125-04	pH	Cool 4 deg C	PSEG01	O31	13893	05/30/2019	9040C
05/30/2019	Water	K3125-05	pH	Cool 4 deg C	PSEG01	O31	13894	05/30/2019	9040C
05/30/2019	Water	K3125-06	pH	Cool 4 deg C	PSEG01	O31	13895	05/30/2019	9040C
05/30/2019	Water	K3125-07	pH	Cool 4 deg C	PSEG01	O31	13896	05/30/2019	9040C
05/30/2019	Water	K3125-08	pH	Cool 4 deg C	PSEG01	O31	13897	05/30/2019	9040C
05/30/2019	Water	K3125-09	pH	Cool 4 deg C	PSEG01	O31	13898	05/30/2019	9040C
05/30/2019	Water	K3125-10	pH	Cool 4 deg C	PSEG01	O31	13900	05/30/2019	9040C
05/30/2019	Water	K3125-11	pH	Cool 4 deg C	PSEG01	O31	13901	05/30/2019	9040C
05/30/2019	Water	K3125-12	pH	Cool 4 deg C	PSEG01	O31	13902	05/30/2019	9040C
05/30/2019	Water	K3125-13	pH	Cool 4 deg C	PSEG01	O31	13905	05/30/2019	9040C

Date/Time 05.30.19 16:15
 Received by: J1
 Relinquished by: PC

Date/Time 05.30.19 18.40
 Received by: PC
 Relinquished by: J1