



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB103109

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	05/30/2019	18:00
2	CAL2	1	Water	NA	NA	20.1	6.99	05/30/2019	18:01
3	CAL3	1	Water	NA	NA	20.1	10.02	05/30/2019	18:02
4	ICV	1	Water	NA	NA	20.1	7.01	05/30/2019	18:03
5	CCV1	1	Water	NA	NA	20.1	2.01	05/30/2019	18:50
6	K3125-19	1	Solid	20.02	20	20.5	7.13	05/30/2019	18:51
7	K3125-19DUP	1	Solid	20.03	20	20.6	7.14	05/30/2019	19:00
8	K3125-20	1	Solid	20.04	20	20.7	5.61	05/30/2019	19:01
9	K3125-21	1	Solid	20.03	20	22.6	5.75	05/30/2019	19:02
10	K3125-22	1	Solid	20.04	20	20.1	5.92	05/30/2019	19:03
11	CCV2	1	Water	NA	NA	20.2	12.02	05/30/2019	19:05

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH S K3140

WorkList ID : 126776

Date : 05-30-2019 15:38:29

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/30/2019	Solid	K3125-19	pH	Cool 4 deg C	PSEG01	O31	13904	05/30/2019	9045D
05/30/2019	Solid	K3125-20	pH	Cool 4 deg C	PSEG01	O31	13899	05/30/2019	9045D
05/30/2019	Solid	K3125-21	pH	Cool 4 deg C	PSEG01	O31	13903	05/30/2019	9045D
05/30/2019	Solid	K3125-22	pH	Cool 4 deg C	PSEG01	O31	13892	05/30/2019	9045D

Date/Time 05-30-19 16:15
 Received by: JH PC
 Relinquished by: JH

Date/Time 05-30-19 16:40
 Received by: PC
 Relinquished by: JH