

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB103123

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP74163
calibration std. phosphate 1 ppm	WP75151
calibration std. phosphate 0.5 ppm	WP75150
calibration std. phosphate 0.3 ppm	WP75149
calibration std. 0 ppm	WP75146
phosphate ICV std.	WP75153
phosphate CCV std.	WP75152
calibration std. phosphate 0.1 ppm	WP75148
calibration std. phosphate 0.05 ppm	WP75147
Combined reagent	WP75156
11 N sulfuric acid	WP71598
Phenolphthalein indicator	WP71599
Sodium hydroxide, 1N	WP71600

Intercept: -0.0006Slope: 0.6459Regression: 0.999873

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	05/31/2019	10:26
2	CAL2	0.05	1	50	50	0.033	05/31/2019	10:27
3	CAL3	0.10	1	50	50	0.065	05/31/2019	10:28
4	CAL4	0.30	1	50	50	0.186	05/31/2019	10:28
5	CAL5	0.50	1	50	50	0.327	05/31/2019	10:29
6	CAL6	1.00	1	50	50	0.645	05/31/2019	10:29



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB103123

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.319	0.495	05/31/2019	10:30
2	ICB		1	50	50	0.000	0.001	05/31/2019	10:30
3	CCV1	0.50	1	50	50	0.321	0.498	05/31/2019	10:31
4	CCB1		1	50	50	0.000	0.001	05/31/2019	10:31
5	PB120240BL		1	50	50	0.000	0.001	05/31/2019	10:32
6	PB120240BS	0.50	1	50	50	0.331	0.513	05/31/2019	10:32
7	K3015-01		1	50	50	0.008	0.013	05/31/2019	10:34
8	K3015-01DUP		1	50	50	0.007	0.012	05/31/2019	10:34
9	K3015-01MS	0.50	1	50	50	0.312	0.484	05/31/2019	10:35
10	K3015-01MSD	0.50	1	50	50	0.315	0.489	05/31/2019	10:35
11	CCV2	0.50	1	50	50	0.326	0.506	05/31/2019	10:36
12	CCB2		1	50	50	0.000	0.001	05/31/2019	10:37