



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 6/3/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 15:15
In Date: 05/31/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:43
Out Date: 06/01/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-1
Thermometer ID: %SOLIDS-OVEN

QC:LB103131

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K3115-01	MH0AH7	1	1.17	9.63	9.08	93.5
K3115-02	MH0AH8	2	1.13	9.79	8.15	81.1
K3115-03	MH0AH9	3	1.18	9.62	8.65	88.5
K3115-04	MH0AJ0	4	1.12	9.56	9.24	96.2
K3115-05	MH0AJ1	5	1.13	9.66	9.31	95.9
K3115-06	MH0AJ2	6	1.15	9.71	9.26	94.7
K3115-07	MH0AJ3	7	1.18	9.54	8.05	82.2
K3115-08	MH0AJ4	8	1.17	9.74	9.47	96.8
K3115-09	MH0AJ5	9	1.12	9.71	9.42	96.6
K3115-10	MH0AJ6	10	1.12	9.96	9.23	91.7
K3115-11	MH0AJ7	11	1.15	9.91	9.48	95.1
K3115-12	MH0AJ8	12	1.18	9.82	9.49	96.2
K3115-13	MH0AJ9	13	1.17	9.63	9.5	98.5
K3115-14	MH0AK0	14	1.19	9.79	8.54	85.5
K3115-15	MH0AK1	15	1.15	9.87	9.61	97.0
K3115-16	MH0AK3	16	1.15	9.84	9.54	96.5
K3115-17	MH0AK4	17	1.15	9.61	9.2	95.2
K3115-18	MH0AK5	18	1.1	9.55	9.36	97.8
K3115-19	MH0AK6	19	1.16	9.94	9.64	96.6
K3115-20	MH0AK8	20	1.14	9.75	9.42	96.2
K3115-21	MH0AK8D	21	1.14	9.75	9.42	96.2
K3115-22	MH0AK8S	22	1.14	9.75	9.42	96.2

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$