



PERCENT SOLID

Supervisor: apatel  
Analyst: JIGNESH  
Date: 6/3/2019

OVENTEMP IN Celsius(°C): 108  
Time IN: 16:45  
In Date: 05/31/2019  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 07:48  
Out Date: 06/01/2019  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-1  
Thermometer ID: %SOLIDS-OVEN

QC:LB103142

| Lab ID   | Client Sample ID | Dish# | Dish Wt(g)<br>(A) | Dish +<br>Sample Wt(g)<br>(B) | Dish + Dry<br>Sample Wt(g)<br>(C) | % Solid |
|----------|------------------|-------|-------------------|-------------------------------|-----------------------------------|---------|
| K3116-01 | MH0AK9           | 1     | 1.14              | 9.61                          | 9.17                              | 94.8    |
| K3116-02 | MH0AL0           | 2     | 1.18              | 9.67                          | 9.38                              | 96.6    |
| K3116-03 | MH0AL1           | 3     | 1.14              | 9.87                          | 9.67                              | 97.7    |
| K3116-04 | MH0AL2           | 4     | 1.13              | 9.69                          | 9.44                              | 97.1    |
| K3116-05 | MH0AL3           | 5     | 1.12              | 9.53                          | 9.36                              | 98.0    |
| K3116-06 | MH0AL4           | 6     | 1.13              | 9.68                          | 9.5                               | 97.9    |
| K3116-07 | MH0AL5           | 7     | 1.12              | 9.73                          | 9.4                               | 96.2    |
| K3116-08 | MH0AL6           | 8     | 1.13              | 9.97                          | 9.57                              | 95.5    |
| K3116-09 | MH0AL7           | 9     | 1.12              | 9.72                          | 9.54                              | 97.9    |
| K3116-10 | MH0AL8           | 10    | 1.11              | 9.92                          | 9.63                              | 96.7    |
| K3116-11 | MH0AL9           | 11    | 1.11              | 9.81                          | 9.63                              | 97.9    |
| K3116-12 | MH0AM0           | 12    | 1.12              | 9.73                          | 9.28                              | 94.8    |
| K3116-13 | MH0AM1           | 13    | 1.1               | 9.97                          | 9.77                              | 97.7    |
| K3116-14 | MH0AM2           | 14    | 1.11              | 9.71                          | 9.59                              | 98.6    |
| K3116-15 | MH0AM3           | 15    | 1.11              | 9.96                          | 9.71                              | 97.2    |
| K3116-16 | MH0AM5           | 16    | 1.12              | 9.93                          | 9.66                              | 96.9    |
| K3116-17 | MH0AM6           | 17    | 1.12              | 9.9                           | 9.69                              | 97.6    |
| K3116-18 | MH0AM7           | 18    | 1.1               | 9.89                          | 9.73                              | 98.2    |
| K3116-19 | MH0AM8           | 19    | 1.11              | 9.77                          | 9.64                              | 98.5    |
| K3116-20 | MH0AM9           | 20    | 1.1               | 9.86                          | 9.69                              | 98.1    |
| K3116-21 | MH0AM9D          | 21    | 1.1               | 9.86                          | 9.69                              | 98.1    |
| K3116-22 | MH0AM9S          | 22    | 1.1               | 9.86                          | 9.69                              | 98.1    |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$