



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB103145

Ambient Barometric Pressure (mmHg): 750.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	81.2	05/31/2019	09:40
2	K3093-01		1	8	100.00	>212.0	>212.0	05/31/2019	10:10
3	K3093-02		1	8	100.00	>212.0	>212.0	05/31/2019	10:40
4	K3104-02		1	9	100.00	>212.0	>212.0	05/31/2019	11:05
5	K3104-05		1	7	100.00	>212.0	>212.0	05/31/2019	11:35
6	K3104-08		1	8	100.00	>212.0	>212.0	05/31/2019	11:55
7	K3125-14		1	9	100.00	>212.0	>212.0	05/31/2019	12:20
8	K3125-15		1	7	100.00	>212.0	>212.0	05/31/2019	12:50
9	K3125-16		1	9	100.00	>212.0	>212.0	05/31/2019	13:25
10	K3125-24		1	7	100.00	>212.0	>212.0	05/31/2019	13:50
11	K3125-24DUP		1	7	100.00	>212.0	>212.0	05/31/2019	14:20
12	K3125-25		1	9	100.00	>212.0	>212.0	05/31/2019	14:45
13	K3125-26		1	7	100.00	>212.0	>212.0	05/31/2019	15:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB103145

WorkList Name : FP53119

WorkList ID : 126811

Date : 05-31-2019 09:01:26

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/04/2019	Water	K3093-01	C Flash Point	Cool 4 deg C	PSEG01	K11	13907	05/30/2019	1010A
06/04/2019	Water	K3093-02	C Flash Point	Cool 4 deg C	PSEG01	K11	13908	05/30/2019	1010A
06/04/2019	Water	K3104-02	F Flash Point	Cool 4 deg C	PSEG01	O31	251810	05/30/2019	1010A
06/04/2019	Water	K3104-05	F Flash Point	Cool 4 deg C	PSEG01	O31	SUMP	05/30/2019	1010A
06/04/2019	Water	K3104-08	F Flash Point	Cool 4 deg C	PSEG01	O31	PIT-WATER	05/30/2019	1010A
06/04/2019	Water	K3125-14	E Flash Point	Cool 4 deg C	PSEG01	O31	COMP-1	05/30/2019	1010A
06/04/2019	Water	K3125-15	E Flash Point	Cool 4 deg C	PSEG01	O31	COMP-2	05/30/2019	1010A
06/04/2019	Water	K3125-16	E Flash Point	Cool 4 deg C	PSEG01	O31	COMP-3	05/30/2019	1010A
06/04/2019	Water	K3125-24	C Flash Point	Cool 4 deg C	PSEG01	O31	13904	05/30/2019	1010A
06/04/2019	Water	K3125-25	C Flash Point	Cool 4 deg C	PSEG01	O31	13899	05/30/2019	1010A
06/04/2019	Water	K3125-26	C Flash Point	Cool 4 deg C	PSEG01	O31	13892	05/30/2019	1010A

Date/Time 05/31/19 09:15
 Received by: KS
 Relinquished by: CS

Date/Time 05/31/19 15:25
 Received by: CS
 Relinquished by: KS