

Sample ID	Extended II μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0		14 -		0 7470A	PPB	5/31/2019 17:49	Std
0.2		804 -		0.2 7470A	PPB	5/31/2019 17:51	Std
2.5		11388 -		2.5 7470A	PPB	5/31/2019 17:53	Std
5		21255 -		5 7470A	PPB	5/31/2019 17:55	Std
7.5		33615 -		7.5 7470A	PPB	5/31/2019 17:58	Std
10		43548 -		10 7470A	PPB	5/31/2019 18:00	Std
ICV70	ICV70	17542	3.9958 -	7470A	PPB	5/31/2019 18:04	SMPL
ICB70	ICB70	-21	-0.0105 -	7470A	PPB	5/31/2019 18:06	SMPL
CCV51	CCV51	21429	4.8824 -	7470A	PPB	5/31/2019 18:08	SMPL
CCB51	CCB51	-38	-0.0144 -	7470A	PPB	5/31/2019 18:10	SMPL
CRA	CRA	803	0.1775 -	7470A	PPB	5/31/2019 18:12	SMPL
HighStd	HighStd	43219	9.8529 -	7470A	PPB	5/31/2019 18:14	SMPL
ChkStd	ChkStd	30537	6.96 -	7470A	PPB	5/31/2019 18:17	SMPL
PB120250E PBW		130	0.0239 -	7470A	PPB	5/31/2019 18:19	SMPL
PB120250E LCSW		17511	3.9887 -	7470A	PPB	5/31/2019 18:26	SMPL
K3093-01	13907	22	-0.0007 -	7470A	PPB	5/31/2019 18:28	SMPL
K3093-01D	13907DUP	79	0.0123 -	7470A	PPB	5/31/2019 18:30	SMPL
K3093-01M	13907MS	16478	3.7531 -	7470A	PPB	5/31/2019 18:32	SMPL
K3093-01M	13907MSD	16442	3.7448 -	7470A	PPB	5/31/2019 18:34	SMPL
K3093-02	13908	3817	0.865 -	7470A	PPB	5/31/2019 18:36	SMPL
CCV52	CCV52	21573	4.9153 -	7470A	PPB	5/31/2019 18:38	SMPL
CCB52	CCB52	-54	-0.018 -	7470A	PPB	5/31/2019 18:41	SMPL
K3104-01	251810	321	0.0675 -	7470A	PPB	5/31/2019 18:43	SMPL
K3104-04	SUMP	106	0.0185 -	7470A	PPB	5/31/2019 18:45	SMPL
K3104-07	PIT-WATER	167	0.0324 -	7470A	PPB	5/31/2019 18:47	SMPL
K3125-14	COMP-1	163	0.0315 -	7470A	PPB	5/31/2019 18:49	SMPL
K3125-15	COMP-2	166	0.0322 -	7470A	PPB	5/31/2019 18:51	SMPL
K3125-16	COMP-3	555	0.1209 -	7470A	PPB	5/31/2019 18:53	SMPL
K3125-24	13904	275	0.057 -	7470A	PPB	5/31/2019 18:55	SMPL
K3125-25	13899	539	0.1172 -	7470A	PPB	5/31/2019 18:58	SMPL
K3125-26	13892	10144	2.3082 -	7470A	PPB	5/31/2019 19:00	SMPL
PB120251E PBW		122	0.0221 -	7470A	PPB	5/31/2019 19:02	SMPL
CCV53	CCV53	21405	4.8769 -	7470A	PPB	5/31/2019 19:04	SMPL
CCB53	CCB53	159	0.0306 -	7470A	PPB	5/31/2019 19:06	SMPL
PB120251E LCSW		17326	3.9465 -	7470A	PPB	5/31/2019 19:08	SMPL
K2642-08	OK-02-0529	217	0.0438 -	7470A	PPB	5/31/2019 19:10	SMPL
K2642-08D	OK-02-0529	260	0.0536 -	7470A	PPB	5/31/2019 19:12	SMPL
K2642-08M	OK-02-0529	18350	4.1801 -	7470A	PPB	5/31/2019 19:15	SMPL
K2642-08M	OK-02-0529	18147	4.1338 -	7470A	PPB	5/31/2019 19:17	SMPL
K2648-04	TR-04-0529	200	0.0399 -	7470A	PPB	5/31/2019 19:19	SMPL
K3090-04	TP-4-S	314	0.0659 -	7470A	PPB	5/31/2019 19:21	SMPL
K3090-10	TP-4-D	265	0.0547 -	7470A	PPB	5/31/2019 19:23	SMPL
PB120252E PBW		238	0.0486 -	7470A	PPB	5/31/2019 19:25	SMPL
PB120252E LCSW		17122	3.9 -	7470A	PPB	5/31/2019 19:27	SMPL
CCV54	CCV54	21938	4.9985 -	7470A	PPB	5/31/2019 19:29	SMPL
CCB54	CCB54	58	0.0075 -	7470A	PPB	5/31/2019 19:31	SMPL

K2639-06	NB-07-0530	76	0.0116 -	7470A	PPB	5/31/2019 19:34 SMPL
K2639-06D	NB-07-0530	82	0.013 -	7470A	PPB	5/31/2019 19:36 SMPL
K2639-06N	NB-07-0530	18703	4.2606 -	7470A	PPB	5/31/2019 19:38 SMPL
K2639-06N	NB-07-0530	17454	3.9757 -	7470A	PPB	5/31/2019 19:40 SMPL
K2639-08	NB-07-0530	49	0.0055 -	7470A	PPB	5/31/2019 19:42 SMPL
K3104-11	RBR-251-68	82	0.013 -	7470A	PPB	5/31/2019 19:44 SMPL
K3105-02	101-102	98	0.0166 -	7470A	PPB	5/31/2019 19:46 SMPL
K3105-04	APPLE-TRA	125	0.0228 -	7470A	PPB	5/31/2019 19:48 SMPL
K3105-06	TRIANGLE	87	0.0141 -	7470A	PPB	5/31/2019 19:51 SMPL
K3105-08	8-9-TRACK	100	0.0171 -	7470A	PPB	5/31/2019 19:53 SMPL
CCV55	CCV55	20689	4.7136 -	7470A	PPB	5/31/2019 19:55 SMPL
CCB55	CCB55	71	0.0105 -	7470A	PPB	5/31/2019 19:57 SMPL
K3105-10	66-67-TRA	95	0.016 -	7470A	PPB	5/31/2019 19:59 SMPL
K3124-06	TP-25-S	95	0.016 -	7470A	PPB	5/31/2019 20:01 SMPL
K3124-10	TP-25-D	96	0.0162 -	7470A	PPB	5/31/2019 20:03 SMPL
PB120253E	PBW	-7	-0.0073 -	7470A	PPB	5/31/2019 20:05 SMPL
PB120253E	LCSW	16543	3.7679 -	7470A	PPB	5/31/2019 20:08 SMPL
K3104-02	251810	47	0.005 -	7470A	PPB	5/31/2019 20:10 SMPL
K3104-02D	251810DUI	91	0.015 -	7470A	PPB	5/31/2019 20:12 SMPL
K3104-02N	251810MS	17039	3.881 -	7470A	PPB	5/31/2019 20:14 SMPL
K3104-02N	251810MS	17597	4.0083 -	7470A	PPB	5/31/2019 20:16 SMPL
K3104-05	SUMP	46	0.0048 -	7470A	PPB	5/31/2019 20:18 SMPL
CCV56	CCV56	21074	4.8014 -	7470A	PPB	5/31/2019 20:20 SMPL
CCB56	CCB56	-8	-0.0075 -	7470A	PPB	5/31/2019 20:23 SMPL
K3104-08	PIT-WATER	111	0.0196 -	7470A	PPB	5/31/2019 20:25 SMPL
PB120251T	PB120167E	123	0.0223 -	7470A	PPB	5/31/2019 20:27 SMPL
PB120252T	PB12026TE	214	0.0431 -	7470A	PPB	5/31/2019 20:29 SMPL
PB120253T	PB120209T	261	0.0538 -	7470A	PPB	5/31/2019 20:31 SMPL
K3093-01LX5		3	-0.005 -	7470A	PPB	5/31/2019 20:33 SMPL
K3093-01A		16244	3.6997 -	7470A	PPB	5/31/2019 20:35 SMPL
K2642-08LX5		21	-0.0009 -	7470A	PPB	5/31/2019 20:37 SMPL
K2642-08A		18245	4.1561 -	7470A	PPB	5/31/2019 20:40 SMPL
K2639-06LX5		-15	-0.0091 -	7470A	PPB	5/31/2019 20:42 SMPL
K2639-06A		18083	4.1192 -	7470A	PPB	5/31/2019 20:44 SMPL
CCV57	CCV57	21202	4.8306 -	7470A	PPB	5/31/2019 20:46 SMPL
CCB57	CCB57	-50	-0.0171 -	7470A	PPB	5/31/2019 20:48 SMPL
K3104-02LX5		-54	-0.018 -	7470A	PPB	5/31/2019 20:50 SMPL
K3104-02A		16899	3.8491 -	7470A	PPB	5/31/2019 20:52 SMPL
CCV58	CCV58	20950	4.7732 -	7470A	PPB	5/31/2019 20:58 SMPL
CCB58	CCB58	-8	-0.0075 -	7470A	PPB	5/31/2019 21:00 SMPL