



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 6/3/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 15:15
In Date: 06/02/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 102
Time OUT: 07:35
Out Date: 06/03/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-1
Thermometer ID: %SOLIDS-OVEN

QC:LB103151

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K3122-01	MH0AT2	1	1.14	9.64	9.52	98.6
K3122-02	MH0AT3	2	1.18	9.92	9.68	97.3
K3122-03	MH0AT4	3	1.17	9.85	9.63	97.5
K3122-04	MH0AT5	4	1.16	9.91	9.78	98.5
K3122-05	MH0AT6	5	1.18	9.96	9.45	94.2
K3122-06	MH0AT7	6	1.16	9.88	9.7	97.9
K3122-07	MH0AT8	7	1.18	9.88	9.71	98.0
K3122-08	MH0AT9	8	1.15	9.67	9.47	97.7
K3122-09	MH0AT9D	9	1.15	9.67	9.47	97.7
K3122-10	MH0AT9S	10	1.15	9.67	9.47	97.7
K3122-11	MH0AU0	11	1.14	9.77	8.56	86.0
K3122-12	MH0AU1	12	1.11	9.76	8.68	87.5
K3122-13	MH0AU2	13	1.13	9.67	8.65	88.1
K3122-14	MH0AU3	14	1.19	9.82	8.79	88.1
K3122-15	MH0AU4	15	1.17	9.91	9.12	91.0
K3122-16	MH0AU5	16	1.15	9.54	8.8	91.2
K3122-17	MH0AU6	17	1.19	9.75	8.75	88.3
K3122-18	MH0AU7	18	1.1	9.65	8.68	88.7
K3122-19	MH0AU8	19	1.13	9.64	9.17	94.5
K3122-20	MH0AU9	20	1.13	9.62	8.82	90.6
K3122-21	MH0AV0	21	1.17	9.92	8.38	82.4
K3122-22	MH0AV1	22	1.16	9.71	8.62	87.3

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$