

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

Instrument ID : Konelab

5/31/2019 16:39

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	95.227	0.0	0.103	
ICB1	0.812	0.0	0.004	
CCV1	234.131	0.0	0.247	
CCB1	0.715	0.0	0.004	
PB120202BL	0.561	0.0	0.004	
PB120202BS	96.793	0.0	0.104	
MIDPB120202	235.567	0.0	0.249	
K3070-07	6.056	0.0	0.010	
K3070-07DUP	5.698	0.0	0.009	
K3082-07	0.853	0.0	0.004	
CCV2	233.980	0.0	0.247	
CCB2	0.655	0.0	0.004	

N	12
Mean	75.921
SD	101.9792
CV%	134.32

AS

Aquakem v. 7.2AQ1

Results from time period:

Fri May 31 16:32:55 2019

Fri May 31 16:35:43 2019

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	0.4403	µg/l	5/31/2019 10:31:56	
5.OPPBCN	A	Total CN	P	5.6121	µg/l	5/31/2019 10:31:57	
10PPBCN	A	Total CN	P	10.6538	µg/l	5/31/2019 10:31:58	
50PPBCN	A	Total CN	P	49.9528	µg/l	5/31/2019 10:31:59	
100PPBCN	A	Total CN	P	100.8352	µg/l	5/31/2019 10:32:00	
250PPBCN	A	Total CN	P	245.3748	µg/l	5/31/2019 10:32:01	
500PPBCN	A	Total CN	P	502.1311	µg/l	5/31/2019 10:32:02	
ICV1	S	Total CN	P	95.2273	µg/l	5/31/2019 16:32:55	
ICB1	S	Total CN	P	0.8124	µg/l	5/31/2019 16:32:56	
CCV1	S	Total CN	P	234.1314	µg/l	5/31/2019 16:32:57	
CCB1	S	Total CN	P	0.7152	µg/l	5/31/2019 16:32:58	
PB120202BL	S	Total CN	P	0.561	µg/l	5/31/2019 16:32:59	
PB120202BS	S	Total CN	P	96.7925	µg/l	5/31/2019 16:33:00	
MIDPB120202	S	Total CN	P	235.567	µg/l	5/31/2019 16:33:01	
K3070-07	S	Total CN	P	6.0559	µg/l	5/31/2019 16:33:02	
K3070-07DUP	S	Total CN	P	5.6976	µg/l	5/31/2019 16:33:03	
K3082-07	S	Total CN	P	0.8526	µg/l	5/31/2019 16:33:04	
CCV2	S	Total CN	P	233.9798	µg/l	5/31/2019 16:33:05	
CCB2	S	Total CN	P	0.6553	µg/l	5/31/2019 16:35:43	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

Instrument ID : Konelab

5/31/2019 16:15


Test Total CN

Accepted 5/31/2019 10:34

Factor 959.7
Bias 0.003

Coeff. of det. 0.999866

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	0.4403	0.0000	
2	5.0PPBCN	0.009	5.6121	5.0000	
3	10PPBCN	0.014	10.6538	10.0000	
4	50PPBCN	0.055	49.9528	50.0000	
5	100PPBCN	0.108	100.8352	100.0000	
6	250PPBCN	0.259	245.3748	250.0000	
7	500PPBCN	0.527	502.1311	500.0000	