

Analysis Method: SM2510 BANALYST: AMANDEEPParameter: ConductanceSUPERVISOR REVIEW BY: apatelRun Number: LB103160Meter ID: WC CONDUCTANCE METER-1Thermometer ID: WC PH METER-1

Reagent/Standard	Lot/Log #
conductivity standard	WP73030

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB103160BL		1	22.3	0.96	1.01	06/03/2019	10:03
2	LB103160BS	1413	1	21.2	1296.00	1397.43	06/03/2019	10:10
3	K3145-01		1	11.8	1.02	1.36	06/03/2019	10:15
4	K3145-02		1	12.3	510.00	673.33	06/03/2019	10:24
5	K3145-03		1	13.5	519.00	665.09	06/03/2019	10:35
6	K3145-03DUP		1	13.6	520.00	664.74	06/03/2019	10:42

Note : Conductivity result is corrected to 25°C

Result = $\frac{\text{(Reading)}}{1 + 0.0191 (\text{Temperature} - 25)}$
--

WORKLIST(Hardcopy Internal Chain)

UB103160

WorkList Name : COND060319

WorkList ID : 126877

Date : 06-03-2019 09:41:13

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/05/2019	Water	K3145-01	Conductance	Cool 4 deg C	STER17	Q12	1A-1B	05/20/2019	SM2510 B
06/05/2019	Water	K3145-02	Conductance	Cool 4 deg C	STER17	Q12	2A-2B	05/20/2019	SM2510 B
06/05/2019	Water	K3145-03	Conductance	Cool 4 deg C	STER17	Q12	3A-3B	05/20/2019	SM2510 B

Date/Time 06/03/19 9:50

Received by: ALO

Relinquished by: [Signature]

Date/Time 06/03/19 10:50

Received by: [Signature]

Relinquished by: [Signature]