



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB103169

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	05/31/2019	10:30
2	CAL2	1	Water	NA	NA	20.1	7.00	05/31/2019	10:31
3	CAL3	1	Water	NA	NA	20.1	10.02	05/31/2019	10:32
4	ICV	1	Water	NA	NA	20.1	7.01	05/31/2019	10:34
5	CCV1	1	Water	NA	NA	20.1	2.01	05/31/2019	10:35
6	K3145-01	1	Water	NA	NA	22.6	4.26	05/31/2019	10:36
7	K3145-02	1	Water	NA	NA	22.5	7.29	05/31/2019	10:38
8	K3145-03	1	Water	NA	NA	22.2	7.64	05/31/2019	10:40
9	K3145-03DUP	1	Water	NA	NA	22.3	7.65	05/31/2019	10:41
10	CCV2	1	Water	NA	NA	22.3	12.02	05/31/2019	10:42

18103169

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K3145 WATER

WorkList ID : 126902

Date : 05-31-2019 10:20:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/20/2019	Water	K3145-01 -1A	pH	Cool 4 deg C	STER17	Q12	1A-1B	05/20/2019	SM 4500-H B
05/20/2019	Water	K3145-02 -1A	pH	Cool 4 deg C	STER17	Q12	2A-2B	05/20/2019	SM 4500-H B
05/20/2019	Water	K3145-03 -1A	pH	Cool 4 deg C	STER17	Q12	3A-3B	05/20/2019	SM 4500-H B

Date/Time 05-31-2019 10:25

Received by: Jh

Relinquished by: CA

Date/Time 05-31-2019 11:20

Received by: Cg

Relinquished by: Jh