



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB103170

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	06/03/2019	15:15
2	CAL2	1	Water	NA	NA	20.1	7.01	06/03/2019	15:16
3	CAL3	1	Water	NA	NA	20.1	10.04	06/03/2019	15:17
4	ICV	1	Water	NA	NA	20.2	7.00	06/03/2019	15:18
5	CCV1	1	Water	NA	NA	20.2	2.01	06/03/2019	15:20
6	K3146-01	1	Solid	20.01	20	21.6	7.11	06/03/2019	15:21
7	K3165-01	1	Solid	20.03	20	22.3	5.69	06/03/2019	15:22
8	K3165-01DUP	1	Solid	20.04	20	22.4	5.70	06/03/2019	15:24
9	K3165-02	1	Solid	20.01	20	21.9	6.24	06/03/2019	15:25
10	K3165-03	1	Solid	20.02	20	21.5	6.90	06/03/2019	15:26
11	K3165-04	1	Solid	20.03	20	21.1	8.54	06/03/2019	15:28
12	K3165-07	1	Solid	20.05	20	21.0	5.98	06/03/2019	15:30
13	K3165-09	1	Solid	20.04	20	20.5	7.50	06/03/2019	15:31
14	CCV2	1	Water	NA	NA	20.1	12.02	06/03/2019	15:33

LB103170

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH S K2151

WorkList ID : 126901

Date : 06-03-2019 15:08:09

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/31/2019	Solid	K3146-01	pH	Cool 4 deg C	PSEG01	O31	OILY-DEBRI-DRUM	05/31/2019	9045D
06/03/2019	Solid	K3165-01	pH	Cool 4 deg C	PSEG01	O31	COMP-1	06/03/2019	9045D
06/03/2019	Solid	K3165-02	pH	Cool 4 deg C	PSEG01	O31	13860	06/03/2019	9045D
06/03/2019	Solid	K3165-03	pH	Cool 4 deg C	PSEG01	O31	30268	06/03/2019	9045D
06/03/2019	Solid	K3165-04	pH	Cool 4 deg C	PSEG01	O31	COMP-2	06/03/2019	9045D
06/03/2019	Solid	K3165-07	pH	Cool 4 deg C	PSEG01	O31	51419	06/03/2019	9045D
06/03/2019	Solid	K3165-09	pH	Cool 4 deg C	PSEG01	O31	13873	06/03/2019	9045D

Date/Time 06.03.19 15:10
Received by: Jn
Relinquished by: ca

Date/Time 06.03.19 16:53
Received by: ca
Relinquished by: Jn