

LB103177


7471B

INSTRUMENT ID: CU1

Linear

 μ Abs.:

59233



A= 0.0000e+000

B= 1.6922e-004

C= -3.0760e-002

Rho= 0.9999569

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.031	-0.031	0	0.000	0				
0.2	0.200	0.182	-0.018	1259	0.0 %	1259				
2.5	2.500	2.547	0.047	15231	0.0 %	15231				
5.0	5.000	5.046	0.046	30002	0.0 %	30002				
7.5	7.500	7.463	-0.037	44280	0.0 %	44280				
10.0	10.000	9.993	-0.007	59233	0.0 %	59233				

LB103177 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 04 Jun 2019 09:56:58

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	0	-	0.00007471B	PPB		04 Jun 2019 09:57:08	S	Std
0.2 - 1	50-2	1259	-	0.20007471B	PPB		04 Jun 2019 09:59:25	S	Std
2.5 - 1	52-5	15231	-	2.50007471B	PPB		04 Jun 2019 10:01:42	S	Std
5.0 - 1	55-0	30002	-	5.00007471B	PPB		04 Jun 2019 10:03:59	S	Std
7.5 - 1	57-5	44280	-	7.50007471B	PPB		04 Jun 2019 10:06:16	S	Std
10.0 - 1	510-0	59233	-	10.00007471B	PPB		04 Jun 2019 10:08:34	S	Std
ICV71 - 1	ICV71	24977	4.1960	-7471B	PPB		04 Jun 2019 10:11:36	U	SMPL
ICB71 - 1	ICB71	-130	-0.0528	-7471B	PPB		04 Jun 2019 10:13:53	U	SMPL
CCV59 - 1	CCV59	29995	5.0451	-7471B	PPB		04 Jun 2019 10:16:08	U	SMPL
CCB59 - 1	CCB59	-94	-0.0467	-7471B	PPB		04 Jun 2019 10:18:23	U	SMPL
CRA - 1	CRA	1254	0.1814	-7471B	PPB		04 Jun 2019 10:20:39	U	SMPL
HighStd - 1	HighStd	58694	9.9017	-7471B	PPB		04 Jun 2019 10:22:54	U	SMPL
ChkStd - 1	ChkStd	40996	6.9068	-7471B	PPB		04 Jun 2019 10:25:11	U	SMPL
PB120302BL - 1	PBS	-164	-0.0585	-7471B	PPB		04 Jun 2019 10:27:29	U	SMPL
PB120302BS - 1	LCSS	22178	3.7223	-7471B	PPB		04 Jun 2019 10:29:47	U	SMPL
K2641-13 - 1	CL-02-053119-A	11165	1.8586	-7471B	PPB		04 Jun 2019 10:32:04	U	SMPL
K2641-13DUP - 1	CL-02-053119-ADUP	10900	1.8138	-7471B	PPB		04 Jun 2019 10:34:20	U	SMPL
K2641-13MS - 1	CL-02-053119-AMS	36808	6.1981	-7471B	PPB		04 Jun 2019 10:42:52	U	SMPL
K2641-13MSD - 1	CL-02-053119-AMSD	36031	6.0666	-7471B	PPB		04 Jun 2019 10:45:09	U	SMPL
K2641-15 - 1	CL-02-053119-B	890	0.1199	-7471B	PPB		04 Jun 2019 10:47:26	U	SMPL
CCV60 - 1	CCV60	30146	5.0707	-7471B	PPB		04 Jun 2019 10:49:43	U	SMPL
CCB60 - 1	CCB60	-76	-0.0436	-7471B	PPB		04 Jun 2019 10:51:58	U	SMPL
K2641-17 - 1	CL-02-053119-C	1228	0.1770	-7471B	PPB		04 Jun 2019 10:54:15	U	SMPL
K2641-19 - 1	CL-02-053119-D	2085	0.3221	-7471B	PPB		04 Jun 2019 10:56:31	U	SMPL
K2641-21 - 1	CL-02-053119-E	2544	0.3997	-7471B	PPB		04 Jun 2019 10:58:46	U	SMPL
K3135-01 - 1	ROLL-OFF-COMP	1089	0.1535	-7471B	PPB		04 Jun 2019 11:01:02	U	SMPL
K3136-01 - 1	FESB-24(6.5-7)	2517	0.3952	-7471B	PPB		04 Jun 2019 11:03:18	U	SMPL
K3138-02 - 1	B-11-COMP	28330	4.7634	-7471B	PPB		04 Jun 2019 11:05:35	U	SMPL
K3138-04 - 1	B-10-COMP	971	0.1336	-7471B	PPB		04 Jun 2019 11:07:51	U	SMPL
K3138-06 - 1	B-09-COMP	1431	0.2114	-7471B	PPB		04 Jun 2019 11:10:09	U	SMPL
K3138-08 - 1	B-08-COMP	1997	0.3072	-7471B	PPB		04 Jun 2019 11:12:27	U	SMPL
K3138-10 - 1	B-07-COMP	2007	0.3089	-7471B	PPB		04 Jun 2019 11:14:42	U	SMPL
CCV61 - 1	CCV61	30799	5.1812	-7471B	PPB		04 Jun 2019 11:16:58	U	SMPL
CCB61 - 1	CCB61	-67	-0.0421	-7471B	PPB		04 Jun 2019 11:19:14	U	SMPL
K3139-03 - 1	TP-1A-S	2443	0.3827	-7471B	PPB		04 Jun 2019 11:21:31	U	SMPL
K3139-07 - 1	TP-1A-D	150	-0.0054	-7471B	PPB		04 Jun 2019 11:23:47	U	SMPL
K3150-01 - 1	IRONBOUND-TP	5014	0.8177	-7471B	PPB		04 Jun 2019 11:26:03	U	SMPL
K3165-01 - 1	COMP-1	349	0.0283	-7471B	PPB		04 Jun 2019 11:28:19	U	SMPL
K3165-02 - 1	13860	131	-0.0086	-7471B	PPB		04 Jun 2019 11:30:35	U	SMPL
K3165-03 - 1	30268	285	0.0175	-7471B	PPB		04 Jun 2019 11:32:52	U	SMPL
K3165-04 - 1	COMP-2	1681	0.2537	-7471B	PPB		04 Jun 2019 11:35:08	U	SMPL
K3165-07 - 1	51419	1409	0.2077	-7471B	PPB		04 Jun 2019 11:37:26	U	SMPL
K3165-09 - 1	13873	1008	0.1398	-7471B	PPB		04 Jun 2019 11:39:43	U	SMPL
K2641-13LX5 - 1		2022	0.3114	-7471B	PPB		04 Jun 2019 11:41:59	U	SMPL
CCV62 - 1	CCV62	30676	5.1604	-7471B	PPB		04 Jun 2019 11:44:14	U	SMPL
CCB62 - 1	CCB62	-69	-0.0424	-7471B	PPB		04 Jun 2019 11:46:30	U	SMPL
K2641-13A - 1		32645	5.4936	-7471B	PPB		04 Jun 2019 11:48:47	U	SMPL
CCV63 - 1	CCV63	30578	5.1438	-7471B	PPB		04 Jun 2019 11:51:03	U	SMPL
CCB63 - 1	CCB63	-117	-0.0506	-7471B	PPB		04 Jun 2019 11:53:19	U	SMPL