

LB103179


7470A

INSTRUMENT ID : CV1

Linear

 μ Abs.:

45229



A= 0.0000e+000

B= 2.2267e-004

C= 1.7308e-003

Rho= 0.9999029

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.022	0.022	92	0.000	92				
0.2	0.200	0.203	0.003	904	0.0 %	904				
2.5	2.500	2.501	0.001	11226	0.0 %	11226				
5.0	5.000	5.001	0.001	22451	0.0 %	22451				
7.5	7.500	7.400	-0.100	33224	0.0 %	33224				
10.0	10.000	10.073	0.073	45229	0.0 %	45229				

LB103179 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 04 Jun 2019 12:56:35

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	92	-	0.0000	7470A	PPB	04 Jun 2019 12:57:38	S	Std
0.2 - 1	50.2	904	-	0.2000	7470A	PPB	04 Jun 2019 12:59:46	S	Std
2.5 - 1	52.5	11226	-	2.5000	7470A	PPB	04 Jun 2019 13:01:55	S	Std
5.0 - 1	55.0	22451	-	5.0000	7470A	PPB	04 Jun 2019 13:04:03	S	Std
7.5 - 1	57.5	33224	-	7.5000	7470A	PPB	04 Jun 2019 13:06:10	S	Std
10.0 - 1	510.0	45229	-	10.0000	7470A	PPB	04 Jun 2019 13:13:08	S	Std
ICV72 - 1	ICV72	18775	4.1823	-	7470A	PPB	04 Jun 2019 13:17:17	U	SMPL
ICB72 - 1	ICB72	10	0.0040	-	7470A	PPB	04 Jun 2019 13:19:23	U	SMPL
CCV64 - 1	CCV64	22466	5.0042	-	7470A	PPB	04 Jun 2019 13:21:30	U	SMPL
CCB64 - 1	CCB64	-19	-0.0025	-	7470A	PPB	04 Jun 2019 13:23:36	U	SMPL
CRA - 1	CRA	901	0.2024	-	7470A	PPB	04 Jun 2019 13:25:42	U	SMPL
HighStd - 1	HighStd	44570	9.9261	-	7470A	PPB	04 Jun 2019 13:27:48	U	SMPL
ChkStd - 1	ChkStd	31335	6.9791	-	7470A	PPB	04 Jun 2019 13:29:54	U	SMPL
PB120304BL - 1	PBW	-101	-0.0208	-	7470A	PPB	04 Jun 2019 13:32:03	U	SMPL
PB120304BS - 1	LCSW	17113	3.8123	-	7470A	PPB	04 Jun 2019 13:34:10	U	SMPL
K2641-14 - 1	CL-02-053119-A	-81	-0.0163	-	7470A	PPB	04 Jun 2019 13:36:18	U	SMPL
K2641-14DUP - 1	CL-02-053119-ADUP	-40	-0.0072	-	7470A	PPB	04 Jun 2019 13:38:25	U	SMPL
K2641-14MS - 1	CL-02-053119-AMS	18389	4.0964	-	7470A	PPB	04 Jun 2019 13:40:33	U	SMPL
K2641-14MSD - 1	CL-02-053119-AMSD	17366	3.8686	-	7470A	PPB	04 Jun 2019 13:42:42	U	SMPL
K2641-16 - 1	CL-02-053119-B	-25	-0.0038	-	7470A	PPB	04 Jun 2019 13:44:49	U	SMPL
CCV65 - 1	CCV65	22477	5.0067	-	7470A	PPB	04 Jun 2019 13:46:56	U	SMPL
CCB65 - 1	CCB65	-26	-0.0041	-	7470A	PPB	04 Jun 2019 13:49:02	U	SMPL
K2641-18 - 1	CL-02-053119-C	-12	-0.0009	-	7470A	PPB	04 Jun 2019 13:51:09	U	SMPL
K2641-20 - 1	CL-02-053119-D	-3	0.0011	-	7470A	PPB	04 Jun 2019 13:53:15	U	SMPL
K2641-22 - 1	CL-02-053119-E	0	0.0017	-	7470A	PPB	04 Jun 2019 13:55:21	U	SMPL
K3135-02 - 1	ROLL-OFF-COMP	13	0.0046	-	7470A	PPB	04 Jun 2019 13:57:28	U	SMPL
K3137-01 - 1	291	14	0.0048	-	7470A	PPB	04 Jun 2019 13:59:35	U	SMPL
K3138-11 - 1	B-11-COMP	18	0.0057	-	7470A	PPB	04 Jun 2019 14:01:43	U	SMPL
K3138-12 - 1	B-10-COMP	-3	0.0011	-	7470A	PPB	04 Jun 2019 14:03:51	U	SMPL
K3138-13 - 1	B-09-COMP	-12	-0.0009	-	7470A	PPB	04 Jun 2019 14:05:59	U	SMPL
K3138-14 - 1	B-08-COMP	0	0.0017	-	7470A	PPB	04 Jun 2019 14:08:11	U	SMPL
K3138-15 - 1	B-07-COMP	5	0.0028	-	7470A	PPB	04 Jun 2019 14:10:21	U	SMPL
CCV66 - 1	CCV66	22350	4.9784	-	7470A	PPB	04 Jun 2019 14:12:30	U	SMPL
CCB66 - 1	CCB66	6	0.0031	-	7470A	PPB	04 Jun 2019 14:14:38	U	SMPL
K3139-04 - 1	TP-1A-S	-12	-0.0009	-	7470A	PPB	04 Jun 2019 14:16:45	U	SMPL
K3139-08 - 1	TP-1A-D	35	0.0095	-	7470A	PPB	04 Jun 2019 14:18:52	U	SMPL
K3146-01 - 1	OILY-DEBRI-DRUM	153	0.0358	-	7470A	PPB	04 Jun 2019 14:20:58	U	SMPL
PB120304TB - 1	PB120254TB	38	0.0102	-	7470A	PPB	04 Jun 2019 14:23:05	U	SMPL
K2641-14LX5 - 1		9	0.0037	-	7470A	PPB	04 Jun 2019 14:25:12	U	SMPL
K2641-14A - 1		18399	4.0986	-	7470A	PPB	04 Jun 2019 14:27:20	U	SMPL
CCV67 - 1	CCV67	22311	4.9697	-	7470A	PPB	04 Jun 2019 14:29:29	U	SMPL
CCB67 - 1	CCB67	-53	-0.0101	-	7470A	PPB	04 Jun 2019 14:31:37	U	SMPL