



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB103182

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.6	06/04/2019	12:45
2	K3165-05		1	7	100.00	>212.0	>212.0	06/04/2019	13:10
3	K3165-05DUP		1	7	100.00	>212.0	>212.0	06/04/2019	13:40
4	K3165-06		1	9	100.00	>212.0	>212.0	06/04/2019	14:05
5	K3165-08		1	8	100.00	>212.0	>212.0	06/04/2019	14:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB103182

WorkList Name : FP6419

WorkList ID : 126936

Date : 06-04-2019 10:18:02

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/07/2019	Water	K3165-05	Flash Point	Cool 4 deg C	PSEG01	O31	13869	06/03/2019	1010A
06/07/2019	Water	K3165-06	Flash Point	Cool 4 deg C	PSEG01	O31	13911	06/03/2019	1010A
06/07/2019	Water	K3165-08	Flash Point	Cool 4 deg C	PSEG01	O31	13870	06/03/2019	1010A

Date/Time 06/04/19 12:20
 Received by: KS
 Relinquished by: CP

Date/Time 06/04/19 14:40
 Received by: CP
 Relinquished by: KS