



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB103191

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	06/04/2019	15:50
2	CAL2	1	Water	NA	NA	20.1	7.01	06/04/2019	15:51
3	CAL3	1	Water	NA	NA	20.1	10.03	06/04/2019	15:52
4	ICV	1	Water	NA	NA	20.1	7.01	06/04/2019	16:00
5	CCV1	1	Water	NA	NA	20.1	2.01	06/04/2019	16:01
6	K3187-01	1	Water	NA	NA	13.2	4.56	06/04/2019	16:02
7	K3187-01DUP	1	Water	NA	NA	13.1	4.57	06/04/2019	16:04
8	CCV2	1	Water	NA	NA	20.2	12.02	06/04/2019	16:05

LB 103191

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K3187 WorkList ID : 126944 Date : 06-04-2019 15:40:33

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/03/2019	Water	K3187-01-1	pH	Cool 4 deg C	STER04	Q52	POOL-WATER	06/03/2019	SM 4500-H B

Date/Time 06-04-19 15:45
Received by: Sh
Relinquished by: _____

Date/Time 06-04-19 16:30
Received by: VB
Relinquished by: VB