



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB103192

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	06/04/2019	15:50
2	CAL2	1	Water	NA	NA	20.1	7.01	06/04/2019	15:51
3	CAL3	1	Water	NA	NA	20.1	10.03	06/04/2019	15:52
4	ICV	1	Water	NA	NA	20.1	7.01	06/04/2019	16:00
5	CCV1	1	Water	NA	NA	20.2	2.02	06/04/2019	16:15
6	K3188-01	1	Solid	20.02	20	21.4	4.31	06/04/2019	16:16
7	K3188-01DUP	1	Solid	20.03	20	21.5	4.32	06/04/2019	16:17
8	K3188-02	1	Solid	20.04	20	20.8	8.10	06/04/2019	16:19
9	K3188-03	1	Solid	20.04	20	20.5	4.73	06/04/2019	16:20
10	K3188-04	1	Solid	20.02	20	20.2	4.65	06/04/2019	16:21
11	CCV2	1	Water	NA	NA	20.2	12.02	06/04/2019	16:22

13 103192

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH S K3188

WorkList ID : 126945

Date : 06-04-2019 15:41:01

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/04/2019	Solid	K3188-01	✓ pH	Cool 4 deg C	TULL02	Q32	FK-PS-01-RAW-SAND	06/04/2019	9045D
06/04/2019	Solid	K3188-02	✓ pH	Cool 4 deg C	TULL02	Q32	FK-PS-01-RAW-COMPOSITE	06/04/2019	9045D
06/04/2019	Solid	K3188-03	✓ pH	Cool 4 deg C	TULL02	Q32	FK-PS-01-CLAY-LUMPS	06/04/2019	9045D
06/04/2019	Solid	K3188-04	✓ pH	Cool 4 deg C	TULL02	Q32	FK-PS-01-RAW-SAND-SILT-C	06/04/2019	9045D

Date/Time 06-04-19 16:00
 Received by: Jo
 Relinquished by: CC

Date/Time 06-04-19 17:20
 Received by: CC
 Relinquished by: Jo