



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB103193

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	06/04/2019	15:50
2	CAL2	1	Water	NA	NA	20.1	7.01	06/04/2019	15:51
3	CAL3	1	Water	NA	NA	20.1	10.03	06/04/2019	15:52
4	ICV	1	Water	NA	NA	20.1	7.01	06/04/2019	16:00
5	CCV1	1	Water	NA	NA	20.2	2.01	06/04/2019	17:10
6	K2638-03	1	Solid	20.02	20	21.3	8.85	06/04/2019	17:11
7	K3150-04	1	Solid	20.03	20	21.9	7.96	06/04/2019	17:12
8	K3151-02	1	Solid	20.02	20	21.4	11.01	06/04/2019	17:13
9	K3151-04	1	Solid	20.04	20	21.4	10.95	06/04/2019	17:15
10	K3151-06	1	Solid	20.04	20	21.0	9.90	06/04/2019	17:18
11	K3163-02	1	Solid	20.02	20	21.6	7.75	06/04/2019	17:20
12	K3163-04	1	Solid	20.01	20	21.0	7.89	06/04/2019	17:22
13	K3182-05	1	Solid	20.04	20	21.8	8.95	06/04/2019	17:25
14	K3182-06	1	Solid	20.03	20	21.3	9.30	06/04/2019	17:26
15	K3182-06DUP	1	Solid	20.04	20	21.4	9.31	06/04/2019	17:29
16	CCV2	1	Water	NA	NA	20.1	12.02	06/04/2019	17:30

LB 103193

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K2682

WorkList ID : 126949

Date : 06-04-2019 15:45:36

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/04/2019	Solid	K2638-03	Corrosivity	Cool 4 deg C	PSEG01	O31	2019	06/04/2019	9045D
06/03/2019	Solid	K3150-04	Corrosivity	Cool 4 deg C	PSEG01	K31	IRONBOUND-TP	06/03/2019	9045D
06/04/2019	Solid	K3151-02	Corrosivity	Cool 4 deg C	PSEG05	O31	AU-06-060419-A	06/04/2019	9045D
06/04/2019	Solid	K3151-04	Corrosivity	Cool 4 deg C	PSEG05	O31	AU-06-060419-B	06/04/2019	9045D
06/04/2019	Solid	K3151-06	Corrosivity	Cool 4 deg C	PSEG05	O31	AU-06-060419-C	06/04/2019	9045D
06/04/2019	Solid	K3163-02	Corrosivity	Cool 4 deg C	PSEG05	O31	SU-02-060419-A	06/04/2019	9045D
06/04/2019	Solid	K3163-04	Corrosivity	Cool 4 deg C	PSEG05	O31	SU-02-060419-B	06/04/2019	9045D
06/03/2019	Solid	K3182-05	Corrosivity	Cool 4 deg C	LIRO01	Q21	B-06-COMP	06/03/2019	9045D
06/03/2019	Solid	K3182-06	Corrosivity	Cool 4 deg C	LIRO01	Q21	B-05-COMP	06/03/2019	9045D

Date/Time 06.04.19 16.20
 Received by: Jn
 Relinquished by: Cb

Date/Time 06.04.19 18.20
 Received by: Cb
 Relinquished by: Jy