

LB103195


7471B

INSTRUMENT ID : C01

Linear

 μ Abs. :

62282



A= 0.0000e+000

B= 1.6008e-004

C= -3.9915e-002

Rho= 0.9998778

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.000	-0.000	249	0.000	249				
0.2	0.200	0.196	-0.004	1474	0.0 %	1474				
2.5	2.500	2.493	-0.007	15821	0.0 %	15821				
5.0	5.000	4.964	-0.036	31262	0.0 %	31262				
7.5	7.500	7.617	0.117	47834	0.0 %	47834				
10.0	10.000	9.930	-0.070	62282	0.0 %	62282				

LB103195 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 05 Jun 2019 09:47:56

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S6-0	249	-	0.0000	7471B	PPB	05 Jun 2019 09:53:47	S	Std
0.2 - 1	S6-2	1474	-	0.2000	7471B	PPB	05 Jun 2019 09:56:10	S	Std
2.5 - 1	S2-5	15821	-	2.5000	7471B	PPB	05 Jun 2019 09:58:27	S	Std
5.0 - 1	S5-0	31262	-	5.0000	7471B	PPB	05 Jun 2019 10:00:44	S	Std
7.5 - 1	S7-5	47834	-	7.5000	7471B	PPB	05 Jun 2019 10:03:02	S	Std
10.0 - 1	S10-0	62282	-	10.0000	7471B	PPB	05 Jun 2019 10:05:18	S	Std
ICV74 - 1	ICV74	24664	3.9082	-	7471B	PPB	05 Jun 2019 10:08:03	U	SMPL
ICB74 - 1	ICB74	219	-0.0049	-	7471B	PPB	05 Jun 2019 10:10:20	U	SMPL
CCV72 - 1	CCV72	31717	5.0372	-	7471B	PPB	05 Jun 2019 10:12:35	U	SMPL
CCB72 - 1	CCB72	201	-0.0077	-	7471B	PPB	05 Jun 2019 10:14:51	U	SMPL
CRA - 1	CRA	1446	0.1916	-	7471B	PPB	05 Jun 2019 10:17:07	U	SMPL
HighStd - 1	HighStd	63083	10.0581	-	7471B	PPB	05 Jun 2019 10:19:23	U	SMPL
ChkStd - 1	ChkStd	44853	7.1399	-	7471B	PPB	05 Jun 2019 10:21:39	U	SMPL
PB120332BL - 1	PBS	193	-0.0090	-	7471B	PPB	05 Jun 2019 10:23:56	U	SMPL
PB120332BS - 1	LCSS	23758	3.7632	-	7471B	PPB	05 Jun 2019 10:26:14	U	SMPL
K3151-01 - 1	AU-06-060419-A	2266	0.3228	-	7471B	PPB	05 Jun 2019 10:28:30	U	SMPL
K3151-01DUP - 1	AU-06-060419-ADUP	2367	0.3390	-	7471B	PPB	05 Jun 2019 10:30:47	U	SMPL
K3151-01MS - 1	AU-06-060419-AMS	26614	4.2203	-	7471B	PPB	05 Jun 2019 10:33:04	U	SMPL
K3151-01MSD - 1	AU-06-060419-AMSD	26205	4.1549	-	7471B	PPB	05 Jun 2019 10:35:22	U	SMPL
K3151-03 - 1	AU-06-060419-B	2311	0.3300	-	7471B	PPB	05 Jun 2019 10:37:38	U	SMPL
CCV73 - 1	CCV73	31499	5.0023	-	7471B	PPB	05 Jun 2019 10:39:55	U	SMPL
CCB73 - 1	CCB73	222	-0.0044	-	7471B	PPB	05 Jun 2019 10:42:11	U	SMPL
K3151-05 - 1	AU-06-060419-C	3594	0.5354	-	7471B	PPB	05 Jun 2019 10:44:28	U	SMPL
K3171-01 - 1	FK-SF-03-008-A	6270	0.9638	-	7471B	PPB	05 Jun 2019 10:46:45	U	SMPL
K3171-03 - 1	FK-SF-03-008-B	7851	1.2168	-	7471B	PPB	05 Jun 2019 10:49:03	U	SMPL
K3172-01 - 1	FK-SF-01-005-A	5966	0.9151	-	7471B	PPB	05 Jun 2019 10:51:22	U	SMPL
K3172-03 - 1	FK-SF-01-005-B	9162	1.4267	-	7471B	PPB	05 Jun 2019 10:53:38	U	SMPL
K3173-01 - 1	FK-SF-02-004-A	6935	1.0702	-	7471B	PPB	05 Jun 2019 10:55:54	U	SMPL
K3173-03 - 1	FK-SF-02-004-B	6102	0.9369	-	7471B	PPB	05 Jun 2019 10:58:12	U	SMPL
K3174-01 - 1	FK-BPM-03-008-ABCD	2418	0.3471	-	7471B	PPB	05 Jun 2019 11:00:29	U	SMPL
K3175-01 - 1	FK-BPM-03-007-ABCD	2129	0.3009	-	7471B	PPB	05 Jun 2019 11:02:47	U	SMPL
K3182-02 - 1	B-06-COMP	29996	4.7617	-	7471B	PPB	05 Jun 2019 11:05:03	U	SMPL
CCV74 - 1	CCV74	31411	4.9882	-	7471B	PPB	05 Jun 2019 11:07:19	U	SMPL
CCB74 - 1	CCB74	194	-0.0089	-	7471B	PPB	05 Jun 2019 11:09:36	U	SMPL
K3182-04 - 1	B-05-COMP	15875	2.5013	-	7471B	PPB	05 Jun 2019 11:11:52	U	SMPL
K3183-01 - 1	SP-COMP-1	7923	1.2284	-	7471B	PPB	05 Jun 2019 11:14:08	U	SMPL
K3183-03 - 1	SP-COMP-2	7143	1.1035	-	7471B	PPB	05 Jun 2019 11:16:24	U	SMPL
K3183-05 - 1	TP-1	5153	0.7850	-	7471B	PPB	05 Jun 2019 11:18:40	U	SMPL
K3183-07 - 1	TP-2	11425	1.7889	-	7471B	PPB	05 Jun 2019 11:20:56	U	SMPL
K3183-09 - 1	TP-3	988	0.1182	-	7471B	PPB	05 Jun 2019 11:23:13	U	SMPL
K3183-11 - 1	TP-4	10961	1.7147	-	7471B	PPB	05 Jun 2019 11:25:30	U	SMPL
PB120333BL - 1	PBS	281	0.0051	-	7471B	PPB	05 Jun 2019 11:27:47	U	SMPL
PB120333BS - 1	LCSS	24187	3.8318	-	7471B	PPB	05 Jun 2019 11:30:05	U	SMPL
K2433-01 - 1	FK-BPM-04-002-ABCD	1210	0.1538	-	7471B	PPB	05 Jun 2019 11:32:21	U	SMPL
CCV75 - 1	CCV75	31921	5.0698	-	7471B	PPB	05 Jun 2019 11:34:37	U	SMPL
CCB75 - 1	CCB75	240	-0.0015	-	7471B	PPB	05 Jun 2019 11:36:53	U	SMPL
K2638-02 - 1	2019	1756	0.2412	-	7471B	PPB	05 Jun 2019 11:39:09	U	SMPL
K3163-01 - 1	SU-02-060419-A	1594	0.2152	-	7471B	PPB	05 Jun 2019 11:41:25	U	SMPL
K3163-03 - 1	SU-02-060419-B	2465	0.3547	-	7471B	PPB	05 Jun 2019 11:43:41	U	SMPL
K3186-01 - 1	FK-EF-01-011-ABCD	7301	1.1288	-	7471B	PPB	05 Jun 2019 11:45:57	U	SMPL
K3186-01DUP - 1	FK-EF-01-011-ABCD DUP	6956	1.0736	-	7471B	PPB	05 Jun 2019 11:48:13	U	SMPL
K3186-01MS - 1	FK-EF-01-011-ABCDMS	29849	4.7382	-	7471B	PPB	05 Jun 2019 11:50:30	U	SMPL
K3186-01MSD - 1	FK-EF-01-011-ABCDMSD	31062	4.9323	-	7471B	PPB	05 Jun 2019 11:52:47	U	SMPL
K3151-01LX5 - 1		548	0.0478	-	7471B	PPB	05 Jun 2019 11:55:04	U	SMPL
K3151-01A - 1		27765	4.4046	-	7471B	PPB	05 Jun 2019 11:57:21	U	SMPL
K3186-01LX5 - 1		1395	0.1834	-	7471B	PPB	05 Jun 2019 11:59:37	U	SMPL
CCV76 - 1	CCV76	32165	5.1089	-	7471B	PPB	05 Jun 2019 12:01:53	U	SMPL
CCB76 - 1	CCB76	235	-0.0023	-	7471B	PPB	05 Jun 2019 12:04:09	U	SMPL
K3186-01A - 1		30480	4.8392	-	7471B	PPB	05 Jun 2019 12:06:25	U	SMPL
CCV77 - 1	CCV77	31611	5.0202	-	7471B	PPB	05 Jun 2019 12:08:42	U	SMPL
CCB77 - 1	CCB77	37	-0.0340	-	7471B	PPB	05 Jun 2019 12:10:58	U	SMPL