



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB103201

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	06/05/2019	11:20
2	K2640-02		1	9	100.00	>212.0	>212.0	06/05/2019	11:55
3	K2640-02DUP		1	9	100.00	>212.0	>212.0	06/05/2019	12:20
4	K2640-04		1	8	100.00	>212.0	>212.0	06/05/2019	12:50
5	K2640-06		1	7	100.00	>212.0	>212.0	06/05/2019	13:25
6	K3182-07		1	6	100.00	>212.0	>212.0	06/05/2019	13:58

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB103201

WorkList Name : FP6519

WorkList ID : 126981

Date : 06-05-2019 09:19:18

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/07/2019	Water	K2640-02	Flash Point	Cool 4 deg C	PSEG01	O31	30234	06/04/2019	1010A
06/07/2019	Water	K2640-04	Flash Point	Cool 4 deg C	PSEG01	O31	30194	06/04/2019	1010A
06/07/2019	Water	K2640-06	Flash Point	Cool 4 deg C	PSEG01	O31	21969	06/04/2019	1010A
06/11/2019	Water	K3182-07	Flash Point	Cool 4 deg C	LIRO01	Q21	MW-10	06/04/2019	1010A

Date/Time 06/05/19 10:50
 Received by: KS
 Relinquished by: CP

Date/Time 06/05/19 14:05
 Received by: CP
 Relinquished by: KS