



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 6/6/2019

OVENTEMP IN Celsius(°C): 107
Time IN: 17:10
In Date: 06/05/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:05
Out Date: 06/06/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-1
Thermometer ID: %SOLIDS-OVEN

QC:LB103216

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K3153-01	HR-05-060519-A	1	1.15	9.97	9.17	90.9
K3153-03	HR-05-060519-B	2	1.12	9.88	9.16	91.8
K3153-05	HR-05-060519-C	3	1.12	9.52	8.98	93.6
K3156-01	HD-02-060519	4	1.17	9.68	9.09	93.1
K3178-03	40061	5	1.00	2.00	2.00	100.0
K3192-01	CULVER-CBTC-WC1B-060419	6	1.15	9.91	9.17	91.6
K3192-03	CULVER-CBTC-WC1A-060419	7	1.14	9.56	9.2	95.7
K3196-01	COMP-1	8	1.15	9.95	7.71	74.5
K3198-01	COMP-DPC-81626-FB-06	9	1.00	2.00	2.00	100.0
K3198-02	COMP-DPC-81626-FB-07	10	1.00	2.00	2.00	100.0
K3198-03	COMP-DPC-81626-FB-08	11	1.00	2.00	2.00	100.0
K3204-01	COMP-1	12	1.12	9.73	8.48	85.5
K3211-01	RBR-20002	13	1.19	9.75	7.93	78.7
K3211-03	RBR-200050	14	1.17	9.88	9.4	94.5
K3211-05	RT-5124	15	1.14	9.96	8.06	78.5
K3211-07	RBR-200052	16	1.18	9.69	8.47	85.7
K3216-01	CONCRETE-COMP-1	17	1.00	2.00	2.00	100.0
K3216-03	CONCRETE-COMP-2	18	1.00	2.00	2.00	100.0
K3218-01	RBR-200035	19	1.12	9.86	8.39	83.2
K3218-02	RBR-200035	20	1.12	9.86	8.39	83.2
K3218-04	RBR-200027	21	1.1	9.76	9.00	91.2
K3218-05	RBR-200027	22	1.1	9.76	9.00	91.2
K3218-07	RT-1873	23	1.16	9.66	8.37	84.8
K3218-08	RT-1873	24	1.16	9.66	8.37	84.8

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$