



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB103217

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	06/05/2019	15:30
2	CAL2	1	Water	NA	NA	20.1	7.00	06/05/2019	15:31
3	CAL3	1	Water	NA	NA	20.1	10.04	06/05/2019	15:33
4	ICV	1	Water	NA	NA	20.1	7.01	06/05/2019	15:35
5	CCV1	1	Water	NA	NA	20.1	2.02	06/05/2019	17:20
6	K3153-02	1	Solid	20.03	20	22.3	9.03	06/05/2019	17:21
7	K3153-02DUP	1	Solid	20.04	20	22.4	9.04	06/05/2019	17:22
8	K3153-04	1	Solid	20.02	20	22.2	8.82	06/05/2019	17:24
9	K3153-06	1	Solid	20.04	20	21.9	7.93	06/05/2019	17:25
10	K3156-02	1	Solid	20.04	20	21.6	9.14	06/05/2019	17:26
11	K3192-02	1	Solid	20.02	20	22.3	7.06	06/05/2019	17:28
12	K3211-02	1	Solid	20.03	20	21.5	8.15	06/05/2019	17:30
13	K3211-04	1	Solid	20.05	20	21.0	5.91	06/05/2019	17:31
14	K3211-06	1	Solid	20.04	20	21.1	5.56	06/05/2019	17:33
15	K3211-08	1	Solid	20.02	20	21.2	7.29	06/05/2019	17:35
16	CCV2	1	Water	NA	NA	20.1	12.02	06/05/2019	17:36
17	K3218-03	1	Solid	20.01	20	21.0	6.47	06/05/2019	17:38
18	K3218-06	1	Solid	20.04	20	21.0	7.05	06/05/2019	17:40
19	K3218-09	1	Solid	20.02	20	21.0	6.62	06/05/2019	17:43
20	CCV3	1	Water	NA	NA	20.1	2.02	06/05/2019	17:45

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K3153

WorkList ID : 127024

Date : 06-05-2019 15:25:09

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/05/2019	Solid	K3153-02	Corrosivity	Cool 4 deg C	PSEG05	O31	HR-05-060519-A	06/05/2019	9045D
06/05/2019	Solid	K3153-04	Corrosivity	Cool 4 deg C	PSEG05	O31	HR-05-060519-B	06/05/2019	9045D
06/05/2019	Solid	K3153-06	Corrosivity	Cool 4 deg C	PSEG05	O31	HR-05-060519-C	06/05/2019	9045D
06/05/2019	Solid	K3156-02	Corrosivity	Cool 4 deg C	PSEG05	O31	HD-02-060519	06/05/2019	9045D
06/04/2019	Solid	K3192-02	Corrosivity	Cool 4 deg C	DIST01	E11	CULVER-CBTC-WC1B-06041	06/04/2019	9045D
06/05/2019	Solid	K3211-02	Corrosivity	Cool 4 deg C	PSEG01	O31	RBR-20002	06/05/2019	9045D
06/05/2019	Solid	K3211-04	Corrosivity	Cool 4 deg C	PSEG01	O31	RBR-200050	06/05/2019	9045D
06/05/2019	Solid	K3211-06	Corrosivity	Cool 4 deg C	PSEG01	O31	RT-5124	06/05/2019	9045D
06/05/2019	Solid	K3211-08	Corrosivity	Cool 4 deg C	PSEG01	O31	RBR-200052	06/05/2019	9045D
06/05/2019	Solid	K3218-03	Corrosivity	Cool 4 deg C	PSEG01	O31	RBR-200035	06/05/2019	9045D
06/05/2019	Solid	K3218-06	Corrosivity	Cool 4 deg C	PSEG01	O31	RBR-200027	06/05/2019	9045D
06/05/2019	Solid	K3218-09	Corrosivity	Cool 4 deg C	PSEG01	O31	RT-1873	06/05/2019	9045D

Date/Time 06-05-19 16:10
 Received by: JA
 Relinquished by: JA

Date/Time 06-05-19 18:20
 Received by: JA
 Relinquished by: JA