


LB103226

7470A

INSTRUMENT ID : CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.013	0.013	-82	0.000	-82				
0.2	0.200	0.219	0.019	892	0.0 %	892				
2.5	2.500	2.500	0.000	11676	0.0 %	11676				
5.0	5.000	4.851	-0.149	22792	0.0 %	22792				
7.5	7.500	7.696	0.196	36244	0.0 %	36244				
10.0	10.000	9.920	-0.080	46759	0.0 %	46759				

LB103226 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 06 Jun 2019 10:45:49

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	-82	-	0.0000	7470A	PPB	06 Jun 2019 11:19:34	S	Std
0.2 - 1	50-2	892	-	0.2000	7470A	PPB	06 Jun 2019 11:21:43	S	Std
2.5 - 1	52-5	11676	-	2.5000	7470A	PPB	06 Jun 2019 11:23:50	S	Std
5.0 - 1	55-0	22792	-	5.0000	7470A	PPB	06 Jun 2019 11:25:58	S	Std
7.5 - 1	57-5	36244	-	7.5000	7470A	PPB	06 Jun 2019 11:28:06	S	Std
10.0 - 1	510-0	46759	-	10.0000	7470A	PPB	06 Jun 2019 11:30:13	S	Std
ICV76 - 1	ICV76	17918	3.8203	-	7470A	PPB	06 Jun 2019 11:32:52	U	SMPL
ICB76 - 1	ICB76	-76	0.0145	-	7470A	PPB	06 Jun 2019 11:34:58	U	SMPL
CCV84 - 1	CCV84	22855	4.8644	-	7470A	PPB	06 Jun 2019 11:37:05	U	SMPL
CCB84 - 1	CCB84	-50	0.0200	-	7470A	PPB	06 Jun 2019 11:39:11	U	SMPL
CRA - 1	CRA	828	0.2057	-	7470A	PPB	06 Jun 2019 11:41:17	U	SMPL
HighStd - 1	HighStd	46229	9.8080	-	7470A	PPB	06 Jun 2019 11:43:23	U	SMPL
ChkStd - 1	ChkStd	33510	7.1180	-	7470A	PPB	06 Jun 2019 11:45:30	U	SMPL
PB120384BL - 1	PBW	-131	0.0029	-	7470A	PPB	06 Jun 2019 11:47:38	U	SMPL
PB120384BS - 1	LCSW	18671	3.9795	-	7470A	PPB	06 Jun 2019 11:49:45	U	SMPL
K2638-03 - 1	2019	-4	0.0297	-	7470A	PPB	06 Jun 2019 11:51:53	U	SMPL
K2638-03DUP - 1	2019DUP	46	0.0403	-	7470A	PPB	06 Jun 2019 11:54:00	U	SMPL
K2638-03MS - 1	2019MS	18748	3.9958	-	7470A	PPB	06 Jun 2019 11:56:08	U	SMPL
K2638-03MSD - 1	2019MSD	19883	4.2358	-	7470A	PPB	06 Jun 2019 11:58:17	U	SMPL
K2640-01 - 1	21643	-53	0.0194	-	7470A	PPB	06 Jun 2019 12:00:23	U	SMPL
CCV85 - 1	CCV85	23171	4.9313	-	7470A	PPB	06 Jun 2019 12:02:29	U	SMPL
CCB85 - 1	CCB85	-75	0.0147	-	7470A	PPB	06 Jun 2019 12:04:35	U	SMPL
K3149-01 - 1	FOAM-COMP	-28	0.0247	-	7470A	PPB	06 Jun 2019 12:06:41	U	SMPL
K3150-04 - 1	IRONBOUND-TP	-17	0.0270	-	7470A	PPB	06 Jun 2019 12:08:48	U	SMPL
K3151-02 - 1	AU-06-060419-A	-11	0.0283	-	7470A	PPB	06 Jun 2019 12:10:54	U	SMPL
K3151-04 - 1	AU-06-060419-B	-25	0.0253	-	7470A	PPB	06 Jun 2019 12:13:01	U	SMPL
K3151-06 - 1	AU-06-060419-C	-15	0.0274	-	7470A	PPB	06 Jun 2019 12:15:08	U	SMPL
K3163-02 - 1	SU-02-060419-A	-17	0.0270	-	7470A	PPB	06 Jun 2019 12:17:15	U	SMPL
K3163-04 - 1	SU-02-060419-B	-19	0.0266	-	7470A	PPB	06 Jun 2019 12:19:22	U	SMPL
K3177-01 - 1	NB-CONCRETE-1	-6	0.0293	-	7470A	PPB	06 Jun 2019 12:21:30	U	SMPL
K3177-02 - 1	BW-CONCRETE-1	8	0.0323	-	7470A	PPB	06 Jun 2019 12:23:38	U	SMPL
K3182-05 - 1	B-06-COMP	14	0.0335	-	7470A	PPB	06 Jun 2019 12:25:45	U	SMPL
CCV86 - 1	CCV86	23156	4.9281	-	7470A	PPB	06 Jun 2019 12:27:52	U	SMPL
CCB86 - 1	CCB86	-77	0.0143	-	7470A	PPB	06 Jun 2019 12:29:58	U	SMPL
K3182-06 - 1	B-05-COMP	15	0.0338	-	7470A	PPB	06 Jun 2019 12:32:05	U	SMPL
PB120385BL - 1	PBW	-18	0.0268	-	7470A	PPB	06 Jun 2019 12:34:11	U	SMPL
PB120385BS - 1	LCSW	18732	3.9924	-	7470A	PPB	06 Jun 2019 12:36:18	U	SMPL
K3165-05 - 1	13869	495	0.1353	-	7470A	PPB	06 Jun 2019 12:38:25	U	SMPL
K3165-06 - 1	13911	5432	1.1795	-	7470A	PPB	06 Jun 2019 12:40:32	U	SMPL
K3165-08 - 1	13870	295	0.0930	-	7470A	PPB	06 Jun 2019 12:42:39	U	SMPL
K3219-01 - 1	EFFL-B2-7	9161	1.9681	-	7470A	PPB	06 Jun 2019 12:44:47	U	SMPL
K3219-01DUP - 1	EFFL-B2-7DUP	9212	1.9789	-	7470A	PPB	06 Jun 2019 12:46:55	U	SMPL
K3219-01MS - 1	EFFL-B2-7MS	24932	5.3037	-	7470A	PPB	06 Jun 2019 12:49:04	U	SMPL
K3219-01MSD - 1	EFFL-B2-7MSD	23589	5.0197	-	7470A	PPB	06 Jun 2019 12:51:10	U	SMPL
CCV87 - 1	CCV87	22669	4.8251	-	7470A	PPB	06 Jun 2019 12:53:19	U	SMPL
CCB87 - 1	CCB87	-127	0.0037	-	7470A	PPB	06 Jun 2019 12:55:27	U	SMPL
PB120384TB - 1	PB120318TB	10	0.0327	-	7470A	PPB	06 Jun 2019 12:57:34	U	SMPL
K2638-03LX5 - 1		-65	0.0168	-	7470A	PPB	06 Jun 2019 12:59:40	U	SMPL
K2638-03A - 1		18756	3.9975	-	7470A	PPB	06 Jun 2019 13:01:47	U	SMPL
K3219-01LX5 - 1		1855	0.4229	-	7470A	PPB	06 Jun 2019 13:03:53	U	SMPL
K3219-01A - 1		24833	5.2828	-	7470A	PPB	06 Jun 2019 13:06:01	U	SMPL
CCV88 - 1	CCV88	22793	4.8513	-	7470A	PPB	06 Jun 2019 13:08:08	U	SMPL
CCB88 - 1	CCB88	-126	0.0039	-	7470A	PPB	06 Jun 2019 13:10:17	U	SMPL