



TOTAL Dissolved Solids - SM2540C

Run Number: LB103231

Date: 06/06/2019

SUPERVISOR: apatel

BalanceID: WC SC-3

ANALYST: Aparna

Filter ID: 14466061

TEMP1 IN:	103	06/05/2019	16:00	TEMP OUT:	103	06/06/2019	07:00	OvenID1:	WC Oven-2	Thermo ID1:	WET-OVEN#2
TEMP2 IN:	182	06/06/2019	07:03	TEMP OUT:	182	06/06/2019	09:00	OvenID2:	WC-OVEN-3	Thermo ID2:	WET-OVEN#3
TEMP3 IN:	182	06/06/2019	10:30	TEMP OUT:	182	06/06/2019	12:00	OvenID3:	WC-OVEN-3	Thermo ID3:	WET-OVEN#3

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @180 (±2) °C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @180 (±2) °C (g)	Final Empty Dish+Sample weight after 1.5hr drying @180 (±2) °C (g)	Weight (g)	Result mg/L
1	LB103231BL	LB103231BL	82.5698	82.5698	100	82.5698	82.5698	82.5698	0	0
2	LB103231BS	LB103231BS	90.1235	90.1235	100	90.134	90.134	90.134	0.0105	105
3	K3190-02	BP-TT-MW201D-20190603	75.1041	75.1041	100	75.1269	75.1269	75.1269	0.0228	228
4	K3190-02DUP	BP-TT-MW201D-20190603DUP	79.9878	79.9878	100	80.0099	80.0099	80.0099	0.0221	221
5	K3190-04	BP-TT-MW201D1-20190603	84.7582	84.7582	100	84.7689	84.7689	84.7689	0.0107	107
6	K3190-06	BP-TT-AOC22-MW08-20190603	84.5557	84.5557	100	84.5622	84.5622	84.5622	0.0065	65
7	K3190-07	BP-TT-AOC22-MW09-20190603	83.4444	83.4444	100	83.4885	83.4885	83.4885	0.0441	441
8	K3190-08	BP-TT-AOC22-MW04-20190603	85.1018	85.1018	100	85.1266	85.1266	85.1266	0.0248	248
9	K3190-10	BP-DUP01-20190603	74.8352	74.8352	100	74.8612	74.8612	74.8612	0.026	260
10	K3190-12	BPS1-TT-MW307D-20190604	81.5306	81.5306	100	81.5408	81.5408	81.5408	0.0102	102
11	K3190-13	BPS1-TT-MW306S-20190604	85.466	85.466	100	85.4888	85.4888	85.4888	0.0228	228
12	K3190-14	BPS1-TT-MW304D-20190604	83.3751	83.3751	100	83.3805	83.3805	83.3805	0.0054	54
13	K3190-15	BPS1-TT-MW304S-30190604	63.8526	63.8526	100	63.8675	63.8675	63.8675	0.0149	149
14	K3190-18	BPS1-TT-MW304I2-20190604	78.4852	78.4852	100	78.49	78.49	78.49	0.0048	48
15	K3190-19	BPS1-TT-MW304I1-20190604	76.7424	76.7424	100	76.7491	76.7491	76.7491	0.0067	67
16	K3190-20	BP-TT-AOC22-MW07-20190604	74.4907	74.4907	100	74.508	74.508	74.508	0.0173	173
17	K3190-21	BP-EB01-20190604	79.1553	79.1553	100	79.1555	79.1555	79.1555	0.0002	2
18	K3190-22	BPS1-TT-MW307I-20190604	84.0502	84.0502	100	84.059	84.059	84.059	0.0088	88
19	K3190-22DUP	BPS1-TT-MW307I-20190604DUP	85.6589	85.6589	100	85.668	85.668	85.668	0.0091	91



TOTAL Dissolved Solids - SM2540C

Run Number: LB103231

SUPERVISOR: apatel

ANALYST: Aparna

Date: 06/06/2019

BalanceID: WC SC-3

Filter ID: 14466061

TEMP1 IN:	<u>103</u>	<u>06/05/2019</u>	<u>16:00</u>	TEMP OUT:	<u>103</u>	<u>06/06/2019</u>	<u>07:00</u>	OvenID1:	<u>WC Oven-2</u>	Thermo ID1:	<u>WET-OVEN#2</u>
TEMP2 IN:	<u>182</u>	<u>06/06/2019</u>	<u>07:03</u>	TEMP OUT:	<u>182</u>	<u>06/06/2019</u>	<u>09:00</u>	OvenID2:	<u>WC-OVEN-3</u>	Thermo ID2:	<u>WET-OVEN#3</u>
TEMP3 IN:	<u>182</u>	<u>06/06/2019</u>	<u>10:30</u>	TEMP OUT:	<u>182</u>	<u>06/06/2019</u>	<u>12:00</u>	OvenID3:	<u>WC-OVEN-3</u>	Thermo ID3:	<u>WET-OVEN#3</u>

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @180 (±2) °C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @180 (±2) °C (g)	Final Empty Dish+Sample weight after 1.5hr drying @180 (±2) °C (g)	Weight (g)	Result mg/L

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = 2nd Empty Dish + Sample weight after 1.5 hr drying @180 (±2) °C (g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS060619

WorkList ID : 127026

Date : 06-06-2019 06:51:48

18105221

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/07/2019	Water	K3180-01	TSS	Cool 4 deg C	TULL01	Q32	001M-WILLETTS-PT-BLVD(JUN)	06/04/2019	SM2540 D
06/06/2019	Water	K3180-02	TSS	Cool 4 deg C	TULL01	Q32	002M-35TH-AVE(JUN)	06/04/2019	SM2540 D
06/07/2019	Water	K3191-01	TSS	Cool 4 deg C	DIST01	E21	CULVER-CBTC-TP1-060419	06/04/2019	SM2540 D
06/12/2019	Water	K3194-01	TSS	Cool 4 deg C	ARAM01	E31	COMP	06/05/2019	SM2540 D

Date/Time

06/06/19, 09:00

Received by:

AS

Relinquished by:

AS

Date/Time

06/06/19, 10:10

Received by:

AS

Relinquished by:

AS