



## Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB103237

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

| Reagent/Standard | Lot/Log # |
|------------------|-----------|
| p-xylene (ICV)   | W2168     |

| Seq | LabID       | True Value °F | DL | Initial Sample °C | Celsius °C | Result °F | Final Result °F | Anal Date  | Anal Time |
|-----|-------------|---------------|----|-------------------|------------|-----------|-----------------|------------|-----------|
| 1   | ICV         | 81            | 1  | 8                 | 27.00      | 80.6      | 80.6            | 06/06/2019 | 13:05     |
| 2   | K3191-01    |               | 1  | 7                 | 100.00     | >212.0    | >212.0          | 06/06/2019 | 13:35     |
| 3   | K3191-01DUP |               | 1  | 9                 | 100.00     | >212.0    | >212.0          | 06/06/2019 | 14:10     |
| 4   | K3196-02    |               | 1  | 7                 | 100.00     | >212.0    | >212.0          | 06/06/2019 | 14:40     |
| 5   | K3204-02    |               | 1  | 8                 | 100.00     | >212.0    | >212.0          | 06/06/2019 | 15:05     |

Result = (Celsius \* 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) \* 0.06

# WORKLIST(Hardcopy Internal Chain)

LB103237

WorkList Name : FP6619

WorkList ID : 127055

Date : 06-06-2019 09:17:40

| Due Date   | Matrix | Sample   | Test        | Preservative | Customer | Storage Location | Customer Sample        | Collect Date | Method |
|------------|--------|----------|-------------|--------------|----------|------------------|------------------------|--------------|--------|
| 06/07/2019 | Water  | K3191-01 | Flash Point | Cool 4 deg C | DIST01   | E21              | CULVER-CBTC-TP1-060419 | 06/04/2019   | 1010A  |
| 06/12/2019 | Water  | K3196-02 | Flash Point | Cool 4 deg C | PSEG01   | O61              | COMP-2                 | 06/05/2019   | 1010A  |
| 06/12/2019 | Water  | K3204-02 | Flash Point | Cool 4 deg C | PSEG01   | O31              | COMP-2                 | 06/05/2019   | 1010A  |

Date/Time 06/06/19 12:30  
 Received by: KS  
 Relinquished by: CP

Date/Time 06/06/19 15:10  
 Received by: CP  
 Relinquished by: KS