

LB103243


ISM02.4

INSTRUMENT ID : CU1

Linear

 μ Abs.:

46350



A= 0.0000e+000

B= 2.1352e-004

C= 1.1355e-002

Rho= 0.9996952

Accept=Unaccepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.022	0.022	48	0.000	48					—
0.05	0.050										
0.20	0.200	0.195	-0.005	860	0.0 %	860					-3
2.50	2.500	2.483	-0.017	11577	0.0 %	11577					-1
5.00	5.000	4.911	-0.089	22948	0.0 %	22948					-2
7.50	7.500	7.681	0.181	35919	0.0 %	35919					2
10.0	10.000	9.908	-0.092	46350	0.0 %	46350					-1

RS
06.06.19

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	48	-		0 ISM02.4	PPB	6/6/2019 16:06	Std
	0.2 S0.2	860	-		0.2 ISM02.4	PPB	6/6/2019 16:08	Std
	2.5 S2.5	11577	-		2.5 ISM02.4	PPB	6/6/2019 16:11	Std
	5 S5.0	22948	-		5 ISM02.4	PPB	6/6/2019 16:13	Std
	7.5 S7.5	35919	-		7.5 ISM02.4	PPB	6/6/2019 16:15	Std
	10 S10.0	46350	-		10 ISM02.4	PPB	6/6/2019 16:17	Std
ICV	ICV	17961	3.8464	-	ISM02.4	PPB	6/6/2019 16:28	SMPL
ICB	ICB	-41	0.0026	-	ISM02.4	PPB	6/6/2019 16:30	SMPL
CCV093	CCV093	22882	4.8971	-	ISM02.4	PPB	6/6/2019 16:33	SMPL
CCB093	CCB093	-28	0.0054	-	ISM02.4	PPB	6/6/2019 16:35	SMPL
PB120400BL	PBW400	38	0.0195	-	ISM02.4	PPB	6/6/2019 16:37	SMPL
K3126-02	MH0AG9	-22	0.0067	-	ISM02.4	PPB	6/6/2019 16:42	SMPL
K3126-03	MH0AH1	203	0.0547	-	ISM02.4	PPB	6/6/2019 16:44	SMPL
K3126-04	MH0AH3	173	0.0483	-	ISM02.4	PPB	6/6/2019 16:46	SMPL
K3126-05	MH0AJ8	82	0.0289	-	ISM02.4	PPB	6/6/2019 16:49	SMPL
K3126-06	MH0AJ9	124	0.0378	-	ISM02.4	PPB	6/6/2019 16:51	SMPL
K3126-07	MH0AK8	35	0.0188	-	ISM02.4	PPB	6/6/2019 16:53	SMPL
K3126-08	MH0AK9	21	0.0158	-	ISM02.4	PPB	6/6/2019 16:55	SMPL
K3126-09	MH0AL9	89	0.0304	-	ISM02.4	PPB	6/6/2019 16:58	SMPL
K3126-10	MH0AM3	482	0.1143	-	ISM02.4	PPB	6/6/2019 17:00	SMPL
K3126-11	MH0AM7	-24	0.0062	-	ISM02.4	PPB	6/6/2019 17:02	SMPL
K3126-12	MH0AN9	12	0.0139	-	ISM02.4	PPB	6/6/2019 17:04	SMPL
K3126-13	MH0AO1	57	0.0235	-	ISM02.4	PPB	6/6/2019 17:07	SMPL
K3126-14	MH0AP2	75	0.0274	-	ISM02.4	PPB	6/6/2019 17:09	SMPL
K3126-15	MH0AQ3	0	0.0114	-	ISM02.4	PPB	6/6/2019 17:11	SMPL
K3126-16	MH0AQ8	38	0.0195	-	ISM02.4	PPB	6/6/2019 17:14	SMPL
K3126-17	MH0AR3	34	0.0186	-	ISM02.4	PPB	6/6/2019 17:16	SMPL
K3126-18	MH0AR5	-2	0.0109	-	ISM02.4	PPB	6/6/2019 17:18	SMPL
K3126-19	MH0AR6	40	0.0199	-	ISM02.4	PPB	6/6/2019 17:20	SMPL
CCV094	CCV094	23803	5.0938	-	ISM02.4	PPB	6/6/2019 17:23	SMPL
CCB094	CCB094	-41	0.0026	-	ISM02.4	PPB	6/6/2019 17:25	SMPL
K3126-20	MH0AR7	287	0.0726	-	ISM02.4	PPB	6/6/2019 17:27	SMPL
K3126-21	MH0AR7D	170	0.0477	-	ISM02.4	PPB	6/6/2019 17:29	SMPL
K3126-22	MH0AR7S	5332	1.1498	-	ISM02.4	PPB	6/6/2019 17:32	SMPL
PB120400TB	LEB400	10	0.0135	-	ISM02.4	PPB	6/6/2019 17:34	SMPL
K3126-01	MH0AF7	158	0.0451	-	ISM02.4	PPB	6/6/2019 17:36	SMPL
CCV095	CCV095	23750	5.0825	-	ISM02.4	PPB	6/6/2019 17:47	SMPL
CCB095	CCB095	-30	0.0049	-	ISM02.4	PPB	6/6/2019 17:49	SMPL