



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 6/7/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 17:15
In Date: 06/06/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:05
Out Date: 06/07/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-1
Thermometer ID: %SOLIDS-OVEN

QC:LB103246

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K3158-01	OK-03-060619	1	1.14	9.87	9.25	92.9
K3160-01	EO-02-060619-A	2	1.12	9.91	9.28	92.8
K3160-03	EO-02-060619-B	3	1.18	9.95	9.3	92.6
K3160-05	EO-02-060619-C	4	1.16	9.73	9.14	93.1
K3212-01	Q610-1B	5	1.14	9.6	7.14	70.9
K3212-03	Q610-2B	6	1.13	9.87	8.34	82.5
K3212-05	Q610-3B	7	1.16	9.63	5.62	52.7
K3212-07	Q610-4B	8	1.11	9.54	6.21	60.5
K3212-09	Q610-5B	9	1.13	9.51	6.37	62.5
K3212-11	Q610-DUP1	10	1.17	9.92	8.46	83.3
K3222-02	TT-007-IDWSO-20190605	11	1.14	9.97	8.88	87.7
K3222-03	TT-008-IDWSO-20190605	12	1.16	9.62	8.1	82.0
K3245-01	DRILLING-MUD	13	1.1	9.77	8.32	83.3
K3246-01	FK-PS-01-047	14	1.11	9.98	8.53	83.7
K3246-02	FK-PS-01-047	15	1.16	9.54	7.76	78.8
K3247-01	FK-GF-02-041-A	16	1.13	9.9	8.81	87.6
K3247-02	FK-GF-02-041-A	17	1.18	9.91	8.85	87.9
K3247-03	FK-GF-02-041-B	18	1.17	9.59	8.6	88.2
K3247-04	FK-GF-02-041-B	19	1.17	9.97	8.9	87.8
K3248-01	FK-GF-02-042-A	20	1.16	9.67	8.62	87.7
K3248-02	FK-GF-02-042-A	21	1.11	9.68	8.61	87.5
K3248-03	FK-GF-02-042-B	22	1.18	9.98	9.11	90.1
K3248-04	FK-GF-02-042-B	23	1.12	9.68	8.66	88.1

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$