



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB103247

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	06/06/2019	17:15
2	CAL2	1	Water	NA	NA	20.1	7.00	06/06/2019	17:16
3	CAL3	1	Water	NA	NA	20.1	10.02	06/06/2019	17:17
4	ICV	1	Water	NA	NA	20.1	7.01	06/06/2019	17:18
5	CCV1	1	Water	NA	NA	20.1	2.02	06/06/2019	17:20
6	K3158-02	1	Solid	20.02	20	23.1	8.60	06/06/2019	17:21
7	K3158-02DUP	1	Solid	20.03	20	23.2	8.61	06/06/2019	17:22
8	K3160-02	1	Solid	20.04	20	22.5	9.51	06/06/2019	17:24
9	K3160-04	1	Solid	20.02	20	22.6	9.04	06/06/2019	17:25
10	K3160-06	1	Solid	20.03	20	22.7	8.97	06/06/2019	17:26
11	CCV2	1	Water	NA	NA	20.2	12.02	06/06/2019	17:30

VB 103245

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity k3160

WorkList ID : 127087

Date : 06-06-2019 15:51:15

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/06/2019	Solid	K3158-02	Corrosivity	Cool 4 deg C	PSEG05	O31	OK-03-060619	06/06/2019	9045D
06/06/2019	Solid	K3160-02	Corrosivity	Cool 4 deg C	PSEG05	O31	EO-02-060619-A	06/06/2019	9045D
06/06/2019	Solid	K3160-04	Corrosivity	Cool 4 deg C	PSEG05	O31	EO-02-060619-B	06/06/2019	9045D
06/06/2019	Solid	K3160-06	Corrosivity	Cool 4 deg C	PSEG05	O31	EO-02-060619-C	06/06/2019	9045D

Date/Time 06-06-19 16:00
 Received by: Jc
 Relinquished by: Jc

Date/Time 06-06-19 18:00
 Received by: Jc
 Relinquished by: Jc