

LB103252


7470A

INSTRUMENT ID : CV1

Linear

 μ Abs.:

43069



A= 0.0000e+000

B= 2.3255e-004

C= 2.7732e-002

Rho= 0.9999404

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.030	0.030	10	0.000	10				
0.2	0.200	0.206	0.006	768	0.0 %	768				
2.5	2.500	2.455	-0.045	10437	0.0 %	10437				
5.0	5.000	5.029	0.029	21506	0.0 %	21506				
7.5	7.500	7.436	-0.064	31859	0.0 %	31859				
10.0	10.000	10.043	0.043	43069	0.0 %	43069				

LB103252 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 07 Jun 2019 09:00:25

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	10	-	0.0000	7470A	PPB	07 Jun 2019 09:21:09	S	Std
0.2 - 1	50-2	768	-	0.2000	7470A	PPB	07 Jun 2019 09:23:23	S	Std
2.5 - 1	52-5	10437	-	2.5000	7470A	PPB	07 Jun 2019 09:25:30	S	Std
5.0 - 1	55-0	21506	-	5.0000	7470A	PPB	07 Jun 2019 09:27:39	S	Std
7.5 - 1	57-5	31859	-	7.5000	7470A	PPB	07 Jun 2019 09:29:46	S	Std
10.0 - 1	510-0	43069	-	10.0000	7470A	PPB	07 Jun 2019 09:34:29	S	Std
ICV79 - 1	ICV79	17560	4.1113	-	7470A	PPB	07 Jun 2019 09:37:06	U	SMPL
ICB79 - 1	ICB79	-69	0.0117	-	7470A	PPB	07 Jun 2019 09:39:12	U	SMPL
CCV96 - 1	CCV96	22232	5.1977	-	7470A	PPB	07 Jun 2019 09:41:19	U	SMPL
CCB96 - 1	CCB96	-109	0.0024	-	7470A	PPB	07 Jun 2019 09:43:25	U	SMPL
CRA - 1	CRA	748	0.2017	-	7470A	PPB	07 Jun 2019 09:45:31	U	SMPL
HighStd - 1	HighStd	41674	9.7190	-	7470A	PPB	07 Jun 2019 09:47:37	U	SMPL
ChkStd - 1	ChkStd	30317	7.0779	-	7470A	PPB	07 Jun 2019 09:49:43	U	SMPL
PB120415BL - 1	PBW	-89	0.0070	-	7470A	PPB	07 Jun 2019 09:51:51	U	SMPL
PB120415BS - 1	LCSW	19189	4.4901	-	7470A	PPB	07 Jun 2019 09:53:59	U	SMPL
K3153-02 - 1	HR-05-060519-A	-52	0.0156	-	7470A	PPB	07 Jun 2019 09:56:07	U	SMPL
K3153-02DUP - 1	HR-05-060519-ADUP	33	0.0354	-	7470A	PPB	07 Jun 2019 09:58:14	U	SMPL
K3153-02MS - 1	HR-05-060519-AMS	17922	4.1955	-	7470A	PPB	07 Jun 2019 10:00:22	U	SMPL
K3153-02MSD - 1	HR-05-060519-AMSD	17807	4.1687	-	7470A	PPB	07 Jun 2019 10:02:30	U	SMPL
K3153-04 - 1	HR-05-060519-B	-60	0.0138	-	7470A	PPB	07 Jun 2019 10:04:37	U	SMPL
CCV97 - 1	CCV97	22070	5.1601	-	7470A	PPB	07 Jun 2019 10:06:44	U	SMPL
CCB97 - 1	CCB97	-34	0.0198	-	7470A	PPB	07 Jun 2019 10:08:50	U	SMPL
K3153-06 - 1	HR-05-060519-C	-21	0.0228	-	7470A	PPB	07 Jun 2019 10:10:57	U	SMPL
K3156-02 - 1	HD-02-060519	-7	0.0261	-	7470A	PPB	07 Jun 2019 10:13:04	U	SMPL
K3192-02 - 1	CULVER-CBTC-WC1B-060419	-7	0.0261	-	7470A	PPB	07 Jun 2019 10:15:10	U	SMPL
K3196-01 - 1	COMP-1	10	0.0301	-	7470A	PPB	07 Jun 2019 10:17:17	U	SMPL
K3198-01 - 1	COMP-DPC-81626-FB-06	0	0.0277	-	7470A	PPB	07 Jun 2019 10:19:25	U	SMPL
K3198-02 - 1	COMP-DPC-81626-FB-07	-19	0.0233	-	7470A	PPB	07 Jun 2019 10:21:32	U	SMPL
K3198-03 - 1	COMP-DPC-81626-FB-08	-4	0.0268	-	7470A	PPB	07 Jun 2019 10:23:40	U	SMPL
K3204-01 - 1	COMP-1	8	0.0296	-	7470A	PPB	07 Jun 2019 10:25:48	U	SMPL
K3211-02 - 1	RBR-20002	-10	0.0254	-	7470A	PPB	07 Jun 2019 10:27:57	U	SMPL
K3211-04 - 1	RBR-200050	-2	0.0273	-	7470A	PPB	07 Jun 2019 10:30:04	U	SMPL
CCV98 - 1	CCV98	22527	5.2664	-	7470A	PPB	07 Jun 2019 10:32:10	U	SMPL
CCB98 - 1	CCB98	-10	0.0254	-	7470A	PPB	07 Jun 2019 10:34:17	U	SMPL
K3211-06 - 1	RT-5124	82	0.0468	-	7470A	PPB	07 Jun 2019 10:36:23	U	SMPL
K3211-08 - 1	RBR-200052	-38	0.0189	-	7470A	PPB	07 Jun 2019 10:38:30	U	SMPL
K3218-03 - 1	RBR-200035	-10	0.0254	-	7470A	PPB	07 Jun 2019 10:40:36	U	SMPL
K3218-06 - 1	RBR-200027	-12	0.0249	-	7470A	PPB	07 Jun 2019 10:42:43	U	SMPL
K3218-09 - 1	RT-1873	-40	0.0184	-	7470A	PPB	07 Jun 2019 10:44:50	U	SMPL
PB120416BL - 1	PBW	6	0.0291	-	7470A	PPB	07 Jun 2019 10:46:57	U	SMPL
PB120416BS - 1	LCSW	17987	4.2106	-	7470A	PPB	07 Jun 2019 10:49:05	U	SMPL
K2640-02 - 1	30234	-44	0.0175	-	7470A	PPB	07 Jun 2019 10:51:13	U	SMPL
K3196-02 - 1	COMP-2	157	0.0642	-	7470A	PPB	07 Jun 2019 10:53:21	U	SMPL
K3196-02DUP - 1	COMP-2DUP	42	0.0375	-	7470A	PPB	07 Jun 2019 10:55:28	U	SMPL
CCV99 - 1	CCV99	22621	5.2882	-	7470A	PPB	07 Jun 2019 10:57:34	U	SMPL
CCB99 - 1	CCB99	-61	0.0135	-	7470A	PPB	07 Jun 2019 10:59:41	U	SMPL
K3196-02MS - 1	COMP-2MS	18485	4.3264	-	7470A	PPB	07 Jun 2019 11:01:48	U	SMPL
K3196-02MSD - 1	COMP-2MSD	18454	4.3192	-	7470A	PPB	07 Jun 2019 11:03:54	U	SMPL
K3204-02 - 1	COMP-2	-71	0.0112	-	7470A	PPB	07 Jun 2019 11:06:01	U	SMPL
PB120417BL - 1	PBW	-23	0.0224	-	7470A	PPB	07 Jun 2019 11:08:08	U	SMPL
PB120417BS - 1	LCSW	18696	4.3755	-	7470A	PPB	07 Jun 2019 11:10:15	U	SMPL
K3190-02 - 1	BP-TT-MW201D-20190603	-40	0.0184	-	7470A	PPB	07 Jun 2019 11:12:23	U	SMPL
K3190-04 - 1	BP-TT-MW201D1-20190603	-55	0.0149	-	7470A	PPB	07 Jun 2019 11:14:31	U	SMPL
K3190-06 - 1	BP-TT-AOC22-MW08-20190603	-44	0.0175	-	7470A	PPB	07 Jun 2019 11:16:39	U	SMPL
K3190-07 - 1	BP-TT-AOC22-MW09-20190603	11	0.0303	-	7470A	PPB	07 Jun 2019 11:18:47	U	SMPL
K3190-08 - 1	BP-TT-AOC22-MW04-20190603	-15	0.0242	-	7470A	PPB	07 Jun 2019 11:20:54	U	SMPL
CCV01 - 1	CCV01	22774	5.3238	-	7470A	PPB	07 Jun 2019 11:23:00	U	SMPL
CCB01 - 1	CCB01	-56	0.0147	-	7470A	PPB	07 Jun 2019 11:25:07	U	SMPL
K3190-10 - 1	BP-DUP01-20190603	-22	0.0226	-	7470A	PPB	07 Jun 2019 11:27:14	U	SMPL
K3190-12 - 1	BPS1-TT-MW307D-20190604	-24	0.0222	-	7470A	PPB	07 Jun 2019 11:29:20	U	SMPL
K3190-13 - 1	BPS1-TT-MW306S-20190604	-53	0.0154	-	7470A	PPB	07 Jun 2019 11:31:27	U	SMPL
K3190-14 - 1	BPS1-TT-MW304D-20190604	-39	0.0187	-	7470A	PPB	07 Jun 2019 11:33:34	U	SMPL
K3190-15 - 1	BPS1-TT-MW304S-30190604	-48	0.0166	-	7470A	PPB	07 Jun 2019 11:35:41	U	SMPL
K3190-15DUP - 1	BPS1-TT-MW304S-30190604DUP	-95	0.0056	-	7470A	PPB	07 Jun 2019 11:37:49	U	SMPL
K3190-16 - 1	K3190-15MS	17693	4.1422	-	7470A	PPB	07 Jun 2019 11:39:56	U	SMPL
K3190-17 - 1	K3190-15MSD	18066	4.2290	-	7470A	PPB	07 Jun 2019 11:42:05	U	SMPL
K3190-18 - 1	BPS1-TT-MW304I2-20190604	-90	0.0068	-	7470A	PPB	07 Jun 2019 11:44:13	U	SMPL
K3190-19 - 1	BPS1-TT-MW304I1-20190604	-31	0.0205	-	7470A	PPB	07 Jun 2019 11:46:20	U	SMPL
CCV02 - 1	CCV02	22374	5.2308	-	7470A	PPB	07 Jun 2019 11:48:28	U	SMPL
CCB02 - 1	CCB02	-69	0.0117	-	7470A	PPB	07 Jun 2019 11:50:35	U	SMPL
K3190-20 - 1	BP-TT-AOC22-MW07-20190604	-29	0.0210	-	7470A	PPB	07 Jun 2019 11:52:42	U	SMPL
K3190-21 - 1	BP-EB01-20190604	-45	0.0173	-	7470A	PPB	07 Jun 2019 11:54:49	U	SMPL
K3190-22 - 1	BPS1-TT-MW307I-20190604	-42	0.0180	-	7470A	PPB	07 Jun 2019 11:56:56	U	SMPL
PB120418BL - 1	PBW	0	0.0277	-	7470A	PPB	07 Jun 2019 11:59:03	U	SMPL
PB120418BS - 1	LCSW	18482	4.3257	-	7470A	PPB	07 Jun 2019 12:01:11	U	SMPL
K3221-02 - 1	BP-TT-AOC22-MW06-20190604	-68	0.0119	-	7470A	PPB	07 Jun 2019 12:03:18	U	SMPL
K3221-02DUP - 1	BP-TT-AOC22-MW06-20190604DUP	-24	0.0222	-	7470A	PPB	07 Jun 2019 12:05:26	U	SMPL
K3221-02MS - 1	BP-TT-AOC22-MW06-20190604MS	18560	4.3438	-	7470A	PPB	07 Jun 2019 12:07:34	U	SMPL
K3221-02MSD - 1	BP-TT-AOC22-MW06-20190604MSD	19457	4.5524	-	7470A	PPB	07 Jun 2019 12:09:42	U	SMPL
K3221-03 - 1	BP-TT-AOC22-MW06-20190604-F	-46	0.0170	-	7470A	PPB	07 Jun 2019 12:11:50	U	SMPL
CCV03 - 1	CCV03	22688	5.3038	-	7470A	PPB	07 Jun 2019 12:13:57	U	SMPL
CCB03 - 1	CCB03	-54	0.0152	-	7470A	PPB	07 Jun 2019 12:16:04	U	SMPL
K3221-04 - 1	BPS1-TT-MW309D-20190605	-34	0.0198	-	7470A	PPB	07 Jun 2019 12:18:11	U	SMPL
K3221-05 - 1	BPS1-TT-MW309D-20190605-F	-44	0.0175	-	7470A	PPB	07 Jun 2019 12:20:18	U	SMPL
K3221-06 - 1	BPS1-TT-MW309S-20190605	-54	0.0152	-	7470A	PPB	07 Jun 2019 12:22:26	U	SMPL
K3221-07 - 1	BPS1-TT-MW309I-20190605	-34	0.0198	-	7470A	PPB	07 Jun 2019 12:24:33	U	SMPL
K3221-08 - 1	BPS1-TT-MW314S-20190605	-32	0.0203	-	7470A	PPB	07 Jun 2019 12:26:40	U	SMPL
K3221-09 - 1	BP-DUP02-20190605	-35	0.0196	-	7470A	PPB	07 Jun 2019 12:28:47	U	SMPL
K3221-10 - 1	BP-DUP02-20190605-F	-23	0.0224	-	7470A	PPB	07 Jun 2019 12:30:55	U	SMPL
K3221-11 - 1	BPS1-TT-MW314I-20190605	-39	0.0187	-	7470A	PPB	07 Jun 2019 12:33:03	U	SMPL
K3221-12 - 1	BPS1-TT-MW314I-20190605-F	-38	0.0189	-	7470A	PPB	07 Jun 2019 12:35:11	U	SMPL
K3221-13 - 1	BPS1-TT-MW301S-20190605	-36	0.0194	-	7470A	PPB	07 Jun 2019 12:37:19	U	SMPL
CCV04 - 1	CCV04	21986	5.1405	-	7470A	PPB	07 Jun 2019 12:39:27	U	SMPL

LB103252 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 07 Jun 2019 09:00:25

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
CCB04 - 1	CCB04	-12	0.0249	-7470A	PPB		07 Jun 2019 12:41:34	U	SMPL
K3221-14 - 1	BPS1-TT-MW301I-20190605	-46	0.0170	-7470A	PPB		07 Jun 2019 12:43:41	U	SMPL
K3221-15 - 1	BPS1-TT-MW301D-20190605	-59	0.0140	-7470A	PPB		07 Jun 2019 12:45:49	U	SMPL
K3221-16 - 1	BPS1-TT-MW313S-20190605	-60	0.0138	-7470A	PPB		07 Jun 2019 12:47:56	U	SMPL
PB120415TB - 1	PB120362TB	-5	0.0266	-7470A	PPB		07 Jun 2019 12:50:03	U	SMPL
PB120414TB - 1	PB120368TB	-16	0.0240	-7470A	PPB		07 Jun 2019 12:52:11	U	SMPL
K3153-02LX5 - 1		-54	0.0152	-7470A	PPB		07 Jun 2019 12:58:22	U	SMPL
K3153-02A - 1		18088	4.2341	-7470A	PPB		07 Jun 2019 13:00:30	U	SMPL
K3196-02LX5 - 1		-44	0.0175	-7470A	PPB		07 Jun 2019 13:02:38	U	SMPL
K3196-02A - 1		18615	4.3566	-7470A	PPB		07 Jun 2019 13:04:46	U	SMPL
K3190-15LX5 - 1		-79	0.0094	-7470A	PPB		07 Jun 2019 13:06:54	U	SMPL
CCB05 - 1 CCV05	CCB05 CCV05	21494	5.0261	-7470A	PPB		07 Jun 2019 13:09:02	U	SMPL
CCB05 - 1	CCB05	-66	0.0124	-7470A	PPB		07 Jun 2019 13:11:10	U	SMPL
K3190-15A - 1		18083	4.2329	-7470A	PPB		07 Jun 2019 13:13:18	U	SMPL
K3221-02LX5 - 1		-61	0.0135	-7470A	PPB		07 Jun 2019 13:15:26	U	SMPL
K3221-02A - 1		18383	4.3027	-7470A	PPB		07 Jun 2019 13:17:34	U	SMPL
CCV06 - 1	CCV06	21310	4.9833	-7470A	PPB		07 Jun 2019 13:19:42	U	SMPL
CCB06 - 1	CCB06	-76	0.0101	-7470A	PPB		07 Jun 2019 13:21:50	U	SMPL

PS 06.07.19