

TOTAL Dissolved Solids - SM2540C

Run Number: LB103253

SUPERVISOR: apatel

ANALYST: Aparna

Date: 06/07/2019

BalanceID: WC SC-3

Filter ID: 14466061

TEMP1 IN:	<u>103</u>	<u>06/06/2019</u>	<u>15:10</u>	TEMP OUT:	<u>103</u>	<u>06/07/2019</u>	<u>07:00</u>	OvenID1:	<u>WC Oven-2</u>	Thermo ID1:	<u>WET-OVEN#2</u>
TEMP2 IN:	<u>182</u>	<u>06/07/2019</u>	<u>07:05</u>	TEMP OUT:	<u>182</u>	<u>06/07/2019</u>	<u>08:40</u>	OvenID2:	<u>WC-OVEN-3</u>	Thermo ID2:	<u>WET-OVEN#3</u>
TEMP3 IN:	<u>182</u>	<u>06/07/2019</u>	<u>10:00</u>	TEMP OUT:	<u>182</u>	<u>06/07/2019</u>	<u>11:45</u>	OvenID3:	<u>WC-OVEN-3</u>	Thermo ID3:	<u>WET-OVEN#3</u>

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @180 (±2) °C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @180 (±2) °C (g)	Final Empty Dish+Sample weight after 1.5hr drying @180 (±2) °C (g)	Weight (g)	Result mg/L
1	LB103253BL	LB103253BL	90.126	90.126	100	90.126	90.126	90.126	0	0
2	LB103253BS	LB103253BS	82.5639	82.5639	100	82.574	82.574	82.574	0.0101	101
3	K3221-02	BP-TT-AOC22-MW06-20190604	104.698	104.698	100	104.7353	104.7353	104.7353	0.0373	373
4	K3221-02DUP	BP-TT-AOC22-MW06-20190604D	84.5546	84.5546	100	84.5924	84.5924	84.5924	0.0378	378
5	K3221-04	BPS1-TT-MW309D-20190605	75.5677	75.5677	100	75.6111	75.6111	75.6111	0.0434	434
6	K3221-06	BPS1-TT-MW309S-20190605	74.8251	74.8251	100	74.852	74.852	74.852	0.0269	269
7	K3221-07	BPS1-TT-MW309I-20190605	85.4673	85.4673	100	85.4936	85.4936	85.4936	0.0263	263
8	K3221-08	BPS1-TT-MW314S-20190605	85.0997	85.0997	100	85.1067	85.1067	85.1067	0.007	70
9	K3221-09	BP-DUP02-20190605	84.7571	84.7571	100	84.7983	84.7983	84.7983	0.0412	412
10	K3221-11	BPS1-TT-MW314I-20190605	63.8443	63.8443	100	63.8681	63.8681	63.8681	0.0238	238
11	K3221-13	BPS1-TT-MW301S-20190605	81.5272	81.5272	100	81.544	81.544	81.544	0.0168	168
12	K3221-14	BPS1-TT-MW301I-20190605	89.1721	89.1721	100	89.183	89.183	89.183	0.0109	109
13	K3221-15	BPS1-TT-MW301D-20190605	83.4443	83.4443	100	83.4699	83.4699	83.4699	0.0256	256
14	K3221-16	BPS1-TT-MW313S-20190605	79.9876	79.9876	100	80.0043	80.0043	80.0043	0.0167	167

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = 2nd Empty Dish + Sample weight after 1.5 hr drying @180 (±2) °C (g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TS060619

WorkList ID : 127082

Date : 06-06-2019 12:48:07

2/16/2021

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/07/2019	Water	K3191-01	TS	Cool 4 deg C	DIST01	E21	CULVER-CBTC-TP1-060419	06/04/2019	SM2540 B

Date/Time 06/07/19 14:45
 Received by: AS
 Relinquished by: AS

Date/Time 06/07/19 15:20
 Received by: AS
 Relinquished by: AS