

OVEN TEMP1: 103 °C TEMP: 105 °C

DATE & TIME IN: 06/06/2019 15:00 OUT: 06/07/2019 07:20

OVEN TEMP2: 105 °C TEMP: 105 °C

DATE & TIME IN: 06/07/2019 09:00 OUT: 06/07/2019 10:40

Run Number: LB103261

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB103261BL	LB103261BL	82.5698	82.5698	100	82.5698	82.5698	82.5698	0	0
2	K3191-01	CULVER-CBTC-TP1-060419	82.4707	82.4707	100	82.8272	82.8272	82.8272	0.3565	3565
3	K3191-01DUP	CULVER-CBTC-TP1-060419 DUP	82.1394	82.1394	100	82.4808	82.4808	82.4808	0.3414	3414

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result mg/L} = ((C - B) / A) * 1000 * 1000$$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TS060619

WorkList ID : 127082

Date : 06-06-2019 12:48:07

2/16/2021

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/07/2019	Water	K3191-01	TS	Cool 4 deg C	DIST01	E21	CULVER-CBTC-TP1-060419	06/04/2019	SM2540 B

Date/Time 06/07/19 14:45
 Received by: AS
 Relinquished by: AS

Date/Time 06/07/19 15:20
 Received by: AS
 Relinquished by: AS