

LB103263


7471B

INSTRUMENT ID : CV1

Linear

 μ Abs. :

64325



A= 0.0000e+000

B= 1.5747e-004

C= 3.2351e-002

Rho= 0.9994523

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.035	0.035	18	0.000	18				
0.2	0.200	0.208	0.008	1115	0.0 %	1115				
2.5	2.500	2.526	0.026	15836	0.0 %	15836				
5.0	5.000	5.017	0.017	31652	0.0 %	31652				
7.5	7.500	7.253	-0.247	45853	0.0 %	45853				
10.0	10.000	10.162	0.162	64325	0.0 %	64325				

LB103263 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 07 Jun 2019 13:36:02

Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50.0	18	-	0.00007471B	PPB	07 Jun 2019 13:37:17	S	Std
0.2 - 1	50.2	1115	-	0.20007471B	PPB	07 Jun 2019 13:39:37	S	Std
2.5 - 1	52.5	15836	-	2.50007471B	PPB	07 Jun 2019 13:41:55	S	Std
5.0 - 1	55.0	31652	-	5.00007471B	PPB	07 Jun 2019 13:44:12	S	Std
7.5 - 1	57.5	45853	-	7.50007471B	PPB	07 Jun 2019 13:46:29	S	Std
10.0 - 1	510.0	64325	-	10.00007471B	PPB	07 Jun 2019 13:52:16	S	Std
ICV80 - 1	ICV80	25260	4.0100	-7471B	PPB	07 Jun 2019 13:56:02	U	SMPL
ICB80 - 1	ICB80	-73	0.0209	-7471B	PPB	07 Jun 2019 13:58:17	U	SMPL
CCV07 - 1	CCV07	32177	5.0992	-7471B	PPB	07 Jun 2019 14:00:33	U	SMPL
CCB07 - 1	CCB07	-88	0.0185	-7471B	PPB	07 Jun 2019 14:02:48	U	SMPL
CRA - 1	CRA	1117	0.2082	-7471B	PPB	07 Jun 2019 14:05:04	U	SMPL
HighStd - 1	HighStd	64601	10.2050	-7471B	PPB	07 Jun 2019 14:07:19	U	SMPL
ChkStd - 1	ChkStd	44532	7.0448	-7471B	PPB	07 Jun 2019 14:09:35	U	SMPL
PB120434BL - 1	PBS	-100	0.0166	-7471B	PPB	07 Jun 2019 14:11:53	U	SMPL
PB120434BS - 1	LCSS	23385	3.7148	-7471B	PPB	07 Jun 2019 14:14:10	U	SMPL
K3158-01 - 1	OK-03-060619	1691	0.2986	-7471B	PPB	07 Jun 2019 14:16:26	U	SMPL
K3158-01DUP - 1	OK-03-060619DUP	1728	0.3045	-7471B	PPB	07 Jun 2019 14:18:44	U	SMPL
K3158-01MS - 1	OK-03-060619MS	25856	4.1039	-7471B	PPB	07 Jun 2019 14:21:01	U	SMPL
K3158-01MSD - 1	OK-03-060619MSD	26712	4.2387	-7471B	PPB	07 Jun 2019 14:23:19	U	SMPL
K3160-01 - 1	EO-02-060619-A	4339	0.7156	-7471B	PPB	07 Jun 2019 14:25:35	U	SMPL
CCV08 - 1	CCV08	32521	5.1534	-7471B	PPB	07 Jun 2019 14:27:51	U	SMPL
CCB08 - 1	CCB08	-112	0.0147	-7471B	PPB	07 Jun 2019 14:30:07	U	SMPL
K3160-03 - 1	EO-02-060619-B	12148	1.9453	-7471B	PPB	07 Jun 2019 14:32:24	U	SMPL
K3160-05 - 1	EO-02-060619-C	5167	0.8460	-7471B	PPB	07 Jun 2019 14:34:39	U	SMPL
K3245-01 - 1	DRILLING-MUD	14	0.0346	-7471B	PPB	07 Jun 2019 14:36:55	U	SMPL
K3246-01 - 1	FK-PS-01-047	2677	0.4539	-7471B	PPB	07 Jun 2019 14:39:11	U	SMPL
K3247-01 - 1	FK-GF-02-041-A	7925	1.2803	-7471B	PPB	07 Jun 2019 14:41:27	U	SMPL
K3247-03 - 1	FK-GF-02-041-B	10230	1.6433	-7471B	PPB	07 Jun 2019 14:43:44	U	SMPL
K3248-01 - 1	FK-GF-02-042-A	9081	1.4623	-7471B	PPB	07 Jun 2019 14:46:01	U	SMPL
K3248-03 - 1	FK-GF-02-042-B	8660	1.3960	-7471B	PPB	07 Jun 2019 14:48:18	U	SMPL
K3250-01 - 1	SCARSDALE-COMPOSITE-1	6620	1.0748	-7471B	PPB	07 Jun 2019 14:50:36	U	SMPL
K3250-02 - 1	CHIMNEY-1	2818	0.4761	-7471B	PPB	07 Jun 2019 14:52:52	U	SMPL
CCV09 - 1	CCV09	31464	4.9870	-7471B	PPB	07 Jun 2019 14:55:10	U	SMPL
CCB09 - 1	CCB09	-119	0.0136	-7471B	PPB	07 Jun 2019 14:57:27	U	SMPL
K3158-01LX5 - 1		311	0.0813	-7471B	PPB	07 Jun 2019 14:59:44	U	SMPL
K3158-01A - 1		26490	4.2037	-7471B	PPB	07 Jun 2019 15:01:59	U	SMPL
CCV10 - 1	CCV10	32592	5.1646	-7471B	PPB	07 Jun 2019 15:04:15	U	SMPL
CCB10 - 1	CCB10	-67	0.0218	-7471B	PPB	07 Jun 2019 15:06:31	U	SMPL