



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB103268

Slope : 97.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/07/2019	11:20
2	CAL2	1	Water	NA	NA	20.2	7.01	06/07/2019	11:21
3	CAL3	1	Water	NA	NA	20.2	10.04	06/07/2019	11:22
4	ICV	1	Water	NA	NA	20.1	6.99	06/07/2019	11:25
5	CCV1	1	Water	NA	NA	20.2	2.01	06/07/2019	11:26
6	K3253-01	1	Water	NA	NA	21.8	4.28	06/07/2019	11:28
7	K3254-01	1	Water	NA	NA	22.8	4.58	06/07/2019	11:30
8	K3254-01DUP	1	Water	NA	NA	22.9	4.59	06/07/2019	11:31
9	CCV2	1	Water	NA	NA	20.1	12.02	06/07/2019	11:32

203268

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph k3153

WorkList ID : 127150

Date : 06-07-2019 11:10:25

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/05/2019	Water	K3253-01	pH	Cool 4 deg C	STER14	H51	JS-10000-SYSTEM	06/05/2019	SM 4500-H B
06/04/2019	Water	K3254-01	pH	Cool 4 deg C	STER15	H11	GILROY-OMEGA-CELL	06/04/2019	SM 4500-H B

Date/Time 06-07-19 11:15
 Received by: Jd
 Relinquished by: Jc

Date/Time 06-07-19 12:00
 Received by: Jc
 Relinquished by: Jd