

LB103273


ISM02.4

INSTRUMENT ID : CV1

Linear

μ Abs.:

5179E



A= 0.0000e+000

B= 1.9418e-004

C= 6.3875e-002

Rho= 0.9996018

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.063	0.063	-2	0.000	-2					—
0.05	0.050										
0.20	0.200	0.236	0.036	884	0.0 %	884					18
2.50	2.500	2.529	0.029	12697	0.0 %	12697					1
5.00	5.000	4.795	-0.205	24365	0.0 %	24365					-4
7.50	7.500	7.455	-0.045	38064	0.0 %	38064					-1
10.0	10.000	10.122	0.122	51798	0.0 %	51798					1

RS 06-07-19

LB103273

INSTRUMENT ID : CV1

K3120-11	MH0AS2	5468	1.1256 -	ISM02.4	PPB	6/7/2019 17:11	SMPL
K3120-12	MH0AS3	4433	0.9247 -	ISM02.4	PPB	6/7/2019 17:13	SMPL
K3120-13	MH0AS4	21266	4.1932 -	ISM02.4	PPB	6/7/2019 17:15	SMPL
K3120-14	MH0AS5	4668	0.9703 -	ISM02.4	PPB	6/7/2019 17:17	SMPL
K3120-15	MH0AS6	2096	0.4709 -	ISM02.4	PPB	6/7/2019 17:20	SMPL
K3120-16	MH0AS7	197	0.1021 -	ISM02.4	PPB	6/7/2019 17:22	SMPL
CCV013	CCV013	24248	4.7722 -	ISM02.4	PPB	6/7/2019 17:28	SMPL
CCB013	CCB013	-72	0.0499 -	ISM02.4	PPB	6/7/2019 17:30	SMPL
K3120-17	MH0AS8	1974	0.4472 -	ISM02.4	PPB	6/7/2019 17:32	SMPL
K3120-18	MH0AS9	7083	1.4392 -	ISM02.4	PPB	6/7/2019 17:34	SMPL
K3120-19	MH0AT0	12514	2.4938 -	ISM02.4	PPB	6/7/2019 17:37	SMPL
K3120-20	MH0AT1	11402	2.2779 -	ISM02.4	PPB	6/7/2019 17:39	SMPL
K3120-21	MH0AT1D	11598	2.3159 -	ISM02.4	PPB	6/7/2019 17:41	SMPL
K3120-22	MH0AT1S	23284	4.585 -	ISM02.4	PPB	6/7/2019 17:44	SMPL
PB120449BL	PBS449	-70	0.0503 -	ISM02.4	PPB	6/7/2019 17:46	SMPL
K3122-01	MH0AT2	9201	1.8505 -	ISM02.4	PPB	6/7/2019 17:48	SMPL
K3122-02	MH0AT3	4734	0.9831 -	ISM02.4	PPB	6/7/2019 17:51	SMPL
K3122-03	MH0AT4	191969	37.3395 -	ISM02.4	PPB	6/7/2019 17:53	SMPL
K3122-04	MH0AT5	278074	54.0589 -	ISM02.4	PPB	6/7/2019 17:55	SMPL
K3122-05	MH0AT6	44651	8.734 -	ISM02.4	PPB	6/7/2019 17:57	SMPL
K3122-06	MH0AT7	14348	2.8499 -	ISM02.4	PPB	6/7/2019 18:00	SMPL
K3122-07	MH0AT8	5432	1.1186 -	ISM02.4	PPB	6/7/2019 18:02	SMPL
K3122-08	MH0AT9	32934	6.4588 -	ISM02.4	PPB	6/7/2019 18:04	SMPL
K3122-09	MH0AT9D	33031	6.4777 -	ISM02.4	PPB	6/7/2019 18:07	SMPL
K3122-10	MH0AT9S	41796	8.1796 -	ISM02.4	PPB	6/7/2019 18:09	SMPL
K3122-11	MH0AU0	17150	3.394 -	ISM02.4	PPB	6/7/2019 18:11	SMPL
K3122-12	MH0AU1	21364	4.2122 -	ISM02.4	PPB	6/7/2019 18:14	SMPL
K3122-13	MH0AU2	9741	1.9553 -	ISM02.4	PPB	6/7/2019 18:16	SMPL
CCV014	CCV014	23741	4.6738 -	ISM02.4	PPB	6/7/2019 18:25	SMPL
CCB014	CCB014	-101	0.0443 -	ISM02.4	PPB	6/7/2019 18:28	SMPL
K3122-14	MH0AU3	5825	1.1949 -	ISM02.4	PPB	6/7/2019 18:30	SMPL
K3122-15	MH0AU4	5419	1.1161 -	ISM02.4	PPB	6/7/2019 18:32	SMPL
K3122-16	MH0AU5	6875	1.3988 -	ISM02.4	PPB	6/7/2019 18:34	SMPL
K3122-17	MH0AU6	4306	0.9 -	ISM02.4	PPB	6/7/2019 18:37	SMPL
K3122-18	MH0AU7	11814	2.3579 -	ISM02.4	PPB	6/7/2019 18:39	SMPL
K3122-19	MH0AU8	7120	1.4464 -	ISM02.4	PPB	6/7/2019 18:41	SMPL
K3122-20	MH0AU9	2233	0.4975 -	ISM02.4	PPB	6/7/2019 18:44	SMPL
K3122-21	MH0AV0	12381	2.468 -	ISM02.4	PPB	6/7/2019 18:46	SMPL
K3122-22	MH0AV1	14681	2.9146 -	ISM02.4	PPB	6/7/2019 18:48	SMPL
K3122-03DLX10		19850	3.9183 -	ISM02.4	PPB	6/7/2019 18:50	SMPL
K3122-04DLX10		28887	5.673 -	ISM02.4	PPB	6/7/2019 18:53	SMPL
CCV015	CCV015	23649	4.6559 -	ISM02.4	PPB	6/7/2019 19:00	SMPL
CCB015	CCB015	-82	0.048 -	ISM02.4	PPB	6/7/2019 19:02	SMPL

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	-2	-		0 ISM02.4	PPB	6/7/2019 15:24	Std
	0.2 S0.2	884	-		0.2 ISM02.4	PPB	6/7/2019 15:26	Std
	2.5 S2.5	12697	-		2.5 ISM02.4	PPB	6/7/2019 15:29	Std
	5 S5.0	24365	-		5 ISM02.4	PPB	6/7/2019 15:31	Std
	7.5 S7.5	38064	-		7.5 ISM02.4	PPB	6/7/2019 15:33	Std
	10 S10.0	51798	-		10 ISM02.4	PPB	6/7/2019 15:36	Std
ICV	ICV	20921	4.1262	-	ISM02.4	PPB	6/7/2019 15:40	SMPL
ICB	ICB	-97	0.045	-	ISM02.4	PPB	6/7/2019 15:42	SMPL
CCV011	CCV011	25238	4.9645	-	ISM02.4	PPB	6/7/2019 15:44	SMPL
CCB011	CCB011	-106	0.0433	-	ISM02.4	PPB	6/7/2019 15:46	SMPL
PB120447BL	PBS447	-26	0.0588	-	ISM02.4	PPB	6/7/2019 15:49	SMPL
K3119-01	MH0AP2	1603	0.3751	-	ISM02.4	PPB	6/7/2019 15:51	SMPL
K3119-02	MH0AP3	5679	1.1666	-	ISM02.4	PPB	6/7/2019 15:53	SMPL
K3119-03	MH0AP4	3565	0.7561	-	ISM02.4	PPB	6/7/2019 15:55	SMPL
K3119-04	MH0AP5	6265	1.2804	-	ISM02.4	PPB	6/7/2019 15:58	SMPL
K3119-05	MH0AP6	9075	1.826	-	ISM02.4	PPB	6/7/2019 16:00	SMPL
K3119-06	MH0AP7	3508	0.745	-	ISM02.4	PPB	6/7/2019 16:02	SMPL
K3119-07	MH0AP8	1770	0.4076	-	ISM02.4	PPB	6/7/2019 16:05	SMPL
K3119-08	MH0AP9	634	0.187	-	ISM02.4	PPB	6/7/2019 16:07	SMPL
K3119-09	MH0AQ0	8659	1.7452	-	ISM02.4	PPB	6/7/2019 16:09	SMPL
K3119-10	MH0AQ1	3517	0.7468	-	ISM02.4	PPB	6/7/2019 16:11	SMPL
K3119-11	MH0AQ2	447	0.1507	-	ISM02.4	PPB	6/7/2019 16:14	SMPL
K3119-12	MH0AQ3	2271	0.5048	-	ISM02.4	PPB	6/7/2019 16:16	SMPL
K3119-13	MH0AQ4	872	0.2332	-	ISM02.4	PPB	6/7/2019 16:18	SMPL
K3119-14	MH0AQ5	4760	0.9881	-	ISM02.4	PPB	6/7/2019 16:20	SMPL
K3119-15	MH0AQ6	786	0.2165	-	ISM02.4	PPB	6/7/2019 16:23	SMPL
K3119-16	MH0AQ7	873	0.2334	-	ISM02.4	PPB	6/7/2019 16:25	SMPL
K3119-17	MH0AQ8	1197	0.2963	-	ISM02.4	PPB	6/7/2019 16:27	SMPL
K3119-18	MH0AQ9	1771	0.4078	-	ISM02.4	PPB	6/7/2019 16:30	SMPL
K3119-19	MH0AR0	3261	0.6971	-	ISM02.4	PPB	6/7/2019 16:32	SMPL
CCV012	CCV012	24267	4.7759	-	ISM02.4	PPB	6/7/2019 16:34	SMPL
CCB012	CCB012	-84	0.0476	-	ISM02.4	PPB	6/7/2019 16:36	SMPL
K3119-20	MH0AR0D	3650	0.7726	-	ISM02.4	PPB	6/7/2019 16:39	SMPL
K3119-21	MH0AR0S	14723	2.9227	-	ISM02.4	PPB	6/7/2019 16:41	SMPL
K3119-22	MH0AR1	688	0.1975	-	ISM02.4	PPB	6/7/2019 16:43	SMPL
PB120448BL	PBS448	-14	0.0612	-	ISM02.4	PPB	6/7/2019 16:46	SMPL
K3120-01	MH0AR2	656	0.1913	-	ISM02.4	PPB	6/7/2019 16:48	SMPL
K3120-02	MH0AR3	1958	0.4441	-	ISM02.4	PPB	6/7/2019 16:50	SMPL
K3120-03	MH0AR4	1093	0.2761	-	ISM02.4	PPB	6/7/2019 16:52	SMPL
K3120-04	MH0AR5	1278	0.312	-	ISM02.4	PPB	6/7/2019 16:55	SMPL
K3120-05	MH0AR6	3119	0.6695	-	ISM02.4	PPB	6/7/2019 16:57	SMPL
K3120-06	MH0AR7	10432	2.0895	-	ISM02.4	PPB	6/7/2019 16:59	SMPL
K3120-07	MH0AR8	894	0.2375	-	ISM02.4	PPB	6/7/2019 17:01	SMPL
K3120-08	MH0AR9	4212	0.8817	-	ISM02.4	PPB	6/7/2019 17:04	SMPL
K3120-09	MH0AS0	1412	0.3381	-	ISM02.4	PPB	6/7/2019 17:06	SMPL
K3120-10	MH0AS1	6227	1.273	-	ISM02.4	PPB	6/7/2019 17:08	SMPL