

LB103299


7470A

INSTRUMENT ID : CV1

Linear ▼

 μ Abs.:

4549E



A= 0.0000e+000

B= 2.2166e-004

C= 6.5504e-003

Rho= 0.9995272

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.050	0.050	198	0.000	198				
0.2	0.200	0.195	-0.005	850	0.0 %	850				
2.5	2.500	2.422	-0.078	10896	0.0 %	10896				
5.0	5.000	5.142	0.142	23168	0.0 %	23168				
7.5	7.500	7.300	-0.200	32903	0.0 %	32903				
10.0	10.000	10.091	0.091	45496	0.0 %	45496				

LB103299 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 10 Jun 2019 10:17:02

Sample ID	Extended ID	µ Abs	Conc	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50-0	198	-	0.00007470A	PPB	10 Jun 2019 10:18:50	S	Std
0.2 - 1	50-2	850	-	0.20007470A	PPB	10 Jun 2019 10:20:58	S	Std
2.5 - 1	52-3	10896	-	2.50007470A	PPB	10 Jun 2019 10:23:06	S	Std
5.0 - 1	55-8	23168	-	5.00007470A	PPB	10 Jun 2019 10:25:14	S	Std
7.5 - 1	57-5	32903	-	7.50007470A	PPB	10 Jun 2019 10:27:21	S	Std
10.0 - 1	510-0	45496	-	10.00007470A	PPB	10 Jun 2019 10:34:52	S	Std
ICV82 - 1	ICV82	17621	3.9124	-7470A	PPB	10 Jun 2019 10:38:19	U	SMPL
ICB82 - 1	ICB82	124	0.0340	-7470A	PPB	10 Jun 2019 10:40:26	U	SMPL
CCV16 - 1	CCV16	23651	5.2490	-7470A	PPB	10 Jun 2019 10:42:31	U	SMPL
CCB16 - 1	CCB16	134	0.0363	-7470A	PPB	10 Jun 2019 10:44:37	U	SMPL
CRA - 1	CRA	823	0.1890	-7470A	PPB	10 Jun 2019 10:46:43	U	SMPL
HighStd - 1	HighStd	46439	10.3002	-7470A	PPB	10 Jun 2019 10:48:50	U	SMPL
ChkStd - 1	ChkStd	31585	7.0076	-7470A	PPB	10 Jun 2019 10:50:56	U	SMPL
PB120474BL - 1	PBW	-33	-0.0008	-7470A	PPB	10 Jun 2019 10:53:04	U	SMPL
PB120474BS - 1	LCSW	17927	3.9802	-7470A	PPB	10 Jun 2019 10:55:11	U	SMPL
K3158-02 - 1	OK-03-060619	15	0.0099	-7470A	PPB	10 Jun 2019 10:57:18	U	SMPL
K3158-02DUP - 1	OK-03-060619DUP	76	0.0234	-7470A	PPB	10 Jun 2019 10:59:26	U	SMPL
K3158-02MS - 1	OK-03-060619MS	18551	4.1185	-7470A	PPB	10 Jun 2019 11:01:33	U	SMPL
K3158-02MSD - 1	OK-03-060619MSD	18447	4.0955	-7470A	PPB	10 Jun 2019 11:03:42	U	SMPL
K3160-02 - 1	EO-02-060619-A	3	0.0072	-7470A	PPB	10 Jun 2019 11:05:48	U	SMPL
CCV17 - 1	CCV17	23636	5.2457	-7470A	PPB	10 Jun 2019 11:07:55	U	SMPL
CCB17 - 1	CCB17	119	0.0329	-7470A	PPB	10 Jun 2019 11:10:01	U	SMPL
K3160-04 - 1	EO-02-060619-B	41	0.0156	-7470A	PPB	10 Jun 2019 11:12:07	U	SMPL
K3160-06 - 1	EO-02-060619-C	50	0.0176	-7470A	PPB	10 Jun 2019 11:14:14	U	SMPL
PB120475BL - 1	PBW	59	0.0196	-7470A	PPB	10 Jun 2019 11:16:21	U	SMPL
PB120475BS - 1	LCSW	18099	4.0184	-7470A	PPB	10 Jun 2019 11:18:27	U	SMPL
K3191-02 - 1	CULVER-CBTC-TP1-060419	60	0.0198	-7470A	PPB	10 Jun 2019 11:20:34	U	SMPL
K3191-02DUP - 1	CULVER-CBTC-TP1-060419 DUP	136	0.0367	-7470A	PPB	10 Jun 2019 11:22:42	U	SMPL
K3191-02MS - 1	CULVER-CBTC-TP1-060419 MS	17435	3.8712	-7470A	PPB	10 Jun 2019 11:24:49	U	SMPL
K3191-02MSD - 1	CULVER-CBTC-TP1-060419 MSD	17653	3.9195	-7470A	PPB	10 Jun 2019 11:26:57	U	SMPL
K3191-07 - 1	CULVER-CBTC-TP1-060419	198	0.0504	-7470A	PPB	10 Jun 2019 11:29:06	U	SMPL
K3236-01 - 1	N48970	42	0.0159	-7470A	PPB	10 Jun 2019 11:31:12	U	SMPL
CCV18 - 1	CCV18	23268	5.1641	-7470A	PPB	10 Jun 2019 11:33:19	U	SMPL
CCB18 - 1	CCB18	-6	0.0052	-7470A	PPB	10 Jun 2019 11:35:26	U	SMPL
PB120474TB - 1	PB120402TB	76	0.0234	-7470A	PPB	10 Jun 2019 11:37:32	U	SMPL
K3158-02LX5 - 1		-20	0.0021	-7470A	PPB	10 Jun 2019 11:39:39	U	SMPL
K3158-02A - 1		18571	4.1230	-7470A	PPB	10 Jun 2019 11:41:47	U	SMPL
K3191-02LX5 - 1		-74	-0.0099	-7470A	PPB	10 Jun 2019 11:43:55	U	SMPL
K3191-02A - 1		17705	3.9310	-7470A	PPB	10 Jun 2019 11:46:02	U	SMPL
CCV19 - 1	CCV19	22732	5.0453	-7470A	PPB	10 Jun 2019 11:48:11	U	SMPL
CCB19 - 1	CCB19	7	0.0081	-7470A	PPB	10 Jun 2019 11:50:19	U	SMPL