

TOTAL SOLIDS - SM2540B

SUPERVISOR: apatel

ANALYST: Aparna

Date: 06/12/2019

OVEN TEMP1: 103 °C TEMP: 105 °C

DATE & TIME IN: 06/10/2019 15:20 OUT: 06/12/2019 07:00

OVEN TEMP2: 105 °C TEMP: 105 °C

DATE & TIME IN: 06/12/2019 08:10 OUT: 06/12/2019 09:45

Run Number: LB103302

BalanceID: WC SC-3

OvenID: WC-OVEN-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB103302BL	LB103302BL	90.1524	90.1524	100	90.1524	90.1524	90.1524	0	0
2	K3265-13	MW-01	79.5261	79.5261	100	85.3064	85.3064	85.3064	5.7803	57803
3	K3265-13DUP	MW-01DUP	74.4918	74.4918	100	80.0739	80.0739	80.0739	5.5821	55821

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

Result mg/L = ((C - B) / A) * 1000 * 1000