

LB103313


245.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

42178



A= 0.0000e+000

B= 2.3941e-004

C= -2.0004e-002

Rho= 0.9994728

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.017	-0.017	11	0.000	11				
0.05	0.050	0.037	-0.013	239	0.0 %	239				
0.2	0.200	0.201	0.001	923	0.0 %	923				
2.5	2.500	2.509	0.009	10565	0.0 %	10565				
5.0	5.000	5.187	0.187	21750	0.0 %	21750				
7.5	7.500	7.255	-0.245	30387	0.0 %	30387				
10.0	10.000	10.078	0.078	42178	0.0 %	42178				

LB103313 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 10 Jun 2019 14:22:11

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	11	-	0.0000	245.1	PPB	10 Jun 2019 14:23:38	S	Std
0.05 - 1	50-05	239	-	0.0500	245.1	PPB	10 Jun 2019 14:25:56	S	Std
0.2 - 1	50-2	923	-	0.2000	245.1	PPB	10 Jun 2019 14:28:13	S	Std
2.5 - 1	52-5	10565	-	2.5000	245.1	PPB	10 Jun 2019 14:30:31	S	Std
5.0 - 1	55-0	21750	-	5.0000	245.1	PPB	10 Jun 2019 14:32:47	S	Std
7.5 - 1	57-5	30387	-	7.5000	245.1	PPB	10 Jun 2019 14:35:03	S	Std
10.0 - 1	510-0	42178	-	10.0000	245.1	PPB	10 Jun 2019 14:37:23	S	Std
ICV84 - 1	ICV84	16182	3.8541	-245.1		PPB	10 Jun 2019 14:40:25	U	SMPL
ICB84 - 1	ICB84	-107	-0.0456	-245.1		PPB	10 Jun 2019 14:42:40	U	SMPL
CCV20 - 1	CCV20	21861	5.2137	-245.1		PPB	10 Jun 2019 14:44:57	U	SMPL
CCB20 - 1	CCB20	-113	-0.0471	-245.1		PPB	10 Jun 2019 14:47:12	U	SMPL
CRA - 1	CRA	246	0.0389	-245.1		PPB	10 Jun 2019 15:02:04	U	SMPL
HighStd - 1	HighStd	43174	10.3162	-245.1		PPB	10 Jun 2019 15:04:20	U	SMPL
ChkStd - 1	ChkStd	29076	6.9410	-245.1		PPB	10 Jun 2019 15:06:36	U	SMPL
PB120492BL - 1	PBW	-139	-0.0533	-245.1		PPB	10 Jun 2019 15:08:53	U	SMPL
PB120492BS - 1	LCSW	18061	4.3040	-245.1		PPB	10 Jun 2019 15:11:10	U	SMPL
K3070-07 - 1	14B-(1-4)-COMP	665	0.1392	-245.1		PPB	10 Jun 2019 15:13:27	U	SMPL
K3070-07DUP - 1	14B-(1-4)-COMP DUP	597	0.1229	-245.1		PPB	10 Jun 2019 15:15:44	U	SMPL
K3070-07MS - 1	14B-(1-4)-COMP MS	14670	3.4921	-245.1		PPB	10 Jun 2019 15:18:02	U	SMPL
K3070-07MSD - 1	14B-(1-4)-COMP MSD	15571	3.7078	-245.1		PPB	10 Jun 2019 15:32:59	U	SMPL
K3082-07 - 1	14B-(1-4)-COMP	564	0.1150	-245.1		PPB	10 Jun 2019 15:35:14	U	SMPL
CCV21 - 1	CCV21	21085	5.0279	-245.1		PPB	10 Jun 2019 15:37:31	U	SMPL
CCB21 - 1	CCB21	-80	-0.0392	-245.1		PPB	10 Jun 2019 15:39:46	U	SMPL
K3182-07 - 1	MW-10	-8	-0.0219	-245.1		PPB	10 Jun 2019 15:42:02	U	SMPL
K3194-01 - 1	COMP	54	-0.0071	-245.1		PPB	10 Jun 2019 15:44:17	U	SMPL
K3265-13 - 1	MW-01	67	-0.0040	-245.1		PPB	10 Jun 2019 15:46:33	U	SMPL
K3070-07LX5 - 1		96	0.0030	-245.1		PPB	10 Jun 2019 15:48:50	U	SMPL
K3070-07A - 1		15332	3.6506	-245.1		PPB	10 Jun 2019 15:51:07	U	SMPL
CCV22 - 1	CCV22	20961	4.9982	-245.1		PPB	10 Jun 2019 15:53:23	U	SMPL
CCB22 - 1	CCB22	-86	-0.0406	-245.1		PPB	10 Jun 2019 15:55:41	U	SMPL