



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB103317

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	06/10/2019	13:05
2	K3265-13		1	9	100.00	>212.0	>212.0	06/10/2019	13:35
3	K3265-13DUP		1	8	100.00	>212.0	>212.0	06/10/2019	13:58

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 103317

WorkList Name : FP61019

WorkList ID : 127194

Date : 06-10-2019 09:18:12

Due Date	Matrix	Sample	Test	Flash Point	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/14/2019	Water	K3265-13	✓	Flash Point	Cool 4 deg C	LIR001	H23	MW-01	06/07/2019	1010A

Date/Time 06/10/19 12:30
 Received by: KS
 Relinquished by: CP

Date/Time 06/10/19 14:05
 Received by: CP
 Relinquished by: KS