



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB103318

Slope : 97.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	06/06/2019	15:20
2	CAL2	1	Water	NA	NA	20.1	7.01	06/06/2019	15:21
3	CAL3	1	Water	NA	NA	20.1	10.02	06/06/2019	15:22
4	ICV	1	Water	NA	NA	20.1	7.00	06/06/2019	15:24
5	CCV1	1	Water	NA	NA	20.2	2.01	06/06/2019	15:25
6	K3236-02	1	Water	NA	NA	21.2	7.34	06/06/2019	15:26
7	K3236-02DUP	1	Water	NA	NA	21.1	7.35	06/06/2019	15:28
8	CCV2	1	Water	NA	NA	20.2	12.01	06/06/2019	15:30

WORKLIST(Hardcopy Internal Chain)

Y8103318

WorkList Name : corrosivity k3236

WorkList ID : 127228

Date : 06-06-2019 15:10:11

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/06/2019	Water	K3236-02 -17	Corrosivity	Cool 4 deg C	PSEG01	O31	N48970	06/06/2019	9040C

Date/Time 06-06-2019 15:15
 Received by: Jg
 Relinquished by: R.

Date/Time 06-06-2019 16:00
 Received by: PC.
 Relinquished by: Jg