



## Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB103319

Slope : 97.5

pH Meter ID : WC PH METER-1

| Calibration Standards          | Chemtech Log# |
|--------------------------------|---------------|
| PH 4 BUFFER SOLUTION           | W2445         |
| buffer solution pH 7 yellow    | W2447         |
| PH 10.01 BUFFER,COLOR CD 475ML | W2391         |
| buffer solution pH 7 yellow    | W2446         |
| pH 2.00 Buffer                 | W2289         |
| Buffer Solution, PH12 (500ml)  | W2467         |

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

| Seq | LabID       | DF | Matrix | Weight (gm) | Volume (ml) | Temperature (°C) | Result (pH) | Anal Date  | Anal Time |
|-----|-------------|----|--------|-------------|-------------|------------------|-------------|------------|-----------|
| 1   | CAL1        | 1  | Water  | NA          | NA          | 20.1             | 3.99        | 06/06/2019 | 15:20     |
| 2   | CAL2        | 1  | Water  | NA          | NA          | 20.1             | 7.01        | 06/06/2019 | 15:21     |
| 3   | CAL3        | 1  | Water  | NA          | NA          | 20.1             | 10.02       | 06/06/2019 | 15:22     |
| 4   | ICV         | 1  | Water  | NA          | NA          | 20.1             | 7.00        | 06/06/2019 | 15:24     |
| 5   | CCV1        | 1  | Water  | NA          | NA          | 20.2             | 2.01        | 06/06/2019 | 15:25     |
| 6   | K3236-01    | 1  | Water  | NA          | NA          | 21.2             | 7.34        | 06/06/2019 | 15:26     |
| 7   | K3236-01DUP | 1  | Water  | NA          | NA          | 21.1             | 7.35        | 06/06/2019 | 15:28     |
| 8   | CCV2        | 1  | Water  | NA          | NA          | 20.2             | 12.01       | 06/06/2019 | 15:30     |

# WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K3236

WorkList ID : 127229

Date : 06-06-2019 15:15:11

| Due Date   | Matrix | Sample       | Test | Preservative | Customer | Storage Location | Customer Sample | Collect Date | Method |
|------------|--------|--------------|------|--------------|----------|------------------|-----------------|--------------|--------|
| 06/06/2019 | Water  | K3236-01 -15 | pH   | Cool 4 deg C | PSEG01   | O31              | N48970          | 06/06/2019   | 9040C  |

Date/Time 06-06-2019 15:15

Received by: Jn

Relinquished by: Re

Date/Time 06-06-2019 16:00

Received by: Pe

Relinquished by: Jo