



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB103320

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/10/2019	15:35
2	CAL2	1	Water	NA	NA	20.2	7.00	06/10/2019	15:36
3	CAL3	1	Water	NA	NA	20.2	10.04	06/10/2019	15:37
4	ICV	1	Water	NA	NA	20.2	7.01	06/10/2019	15:38
5	CCV1	1	Water	NA	NA	20.1	2.02	06/10/2019	15:40
6	K3271-03	1	Solid	20.02	20	21.4	8.36	06/10/2019	15:41
7	K3271-03DUP	1	Solid	20.03	20	21.5	8.37	06/10/2019	15:42
8	CCV2	1	Water	NA	NA	20.2	12.02	06/10/2019	15:44

WORKLIST(Hardcopy Internal Chain)

LB 103320

WorkList Name : PH K3271

WorkList ID : 127230

Date : 06-10-2019 15:28:29

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/07/2019	Solid	K3271-03 - 18	pH	Cool 4 deg C	loul01	G41	BAT-B1-GEO	06/07/2019	9045D

Date/Time 06.10.19 15.30
 Received by: JH
 Relinquished by:

Date/Time 06.10.19 16.35
 Received by: CD
 Relinquished by: JH