



TOTAL Dissolved Solids - SM2540C

Run Number: LB103323

Date: 06/12/2019

SUPERVISOR: apatel

BalanceID: WC SC-3

ANALYST: Aparna

Filter ID: 14466061

TEMP1 IN:	103	06/10/2019	15:40	TEMP OUT:	105	06/12/2019	07:10	OvenID1:	WC Oven-2	Thermo ID1:	WET-OVEN#2
TEMP2 IN:	182	06/12/2019	07:15	TEMP OUT:	182	06/12/2019	09:00	OvenID2:	WC-OVEN-3	Thermo ID2:	WET-OVEN#3
TEMP3 IN:	182	06/12/2019	09:45	TEMP OUT:	182	06/12/2019	11:20	OvenID3:	WC-OVEN-3	Thermo ID3:	WET-OVEN#3

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @180 (±2) °C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @180 (±2) °C (g)	Final Empty Dish+Sample weight after 1.5hr drying @180 (±2) °C (g)	Weight (g)	Result mg/L
1	LB103323BL	LB103323BL	82.5698	82.5698	100	82.5698	82.5698	82.5698	0	0
2	LB103323BS	LB103323BS	90.1526	90.1526	100	90.163	90.163	90.163	0.0104	104
3	K3279-02	BPS1-TT-MW312S-20190607	75.5614	75.5614	100	75.5789	75.5789	75.5789	0.0175	175
4	K3279-02DUP	BPS1-TT-MW312S-20190607DUP	85.0943	85.0943	100	85.112	85.112	85.112	0.0177	177
5	K3279-03	BPS1-TT-MW312I-20190607	78.0813	78.0813	100	78.0997	78.0997	78.0997	0.0184	184
6	K3279-04	BPS1-TT-MW302I2-20190607	77.5803	77.5803	100	77.5996	77.5996	77.5996	0.0193	193
7	K3279-05	BPS1-TT-MW307S-20190607	92.8158	92.8158	100	92.8282	92.8282	92.8282	0.0124	124
8	K3279-06	BPS1-TT-MW302D-20190607	68.8382	68.8382	100	68.8586	68.8586	68.8586	0.0204	204
9	K3279-07	BP-TT-MW203I-20190607	90.0076	90.0076	100	90.0239	90.0239	90.0239	0.0163	163
10	K3279-08	BPS1-TT-MW303I1-20190607	84.5554	84.5554	100	84.5796	84.5796	84.5796	0.0242	242
11	K3279-09	BP-EB03-20190607	74.8373	74.8373	100	74.8375	74.8375	74.8375	0.0002	2
12	K3280-02	BP-FW-MW01-20190608	79.9901	79.9901	100	80.0094	80.0094	80.0094	0.0193	193
13	K3280-02DUP	BP-FW-MW01-20190608DUP	90.1542	90.1542	100	90.174	90.174	90.174	0.0198	198
14	K3280-03	BP-FW-MW02-20190608	81.527	81.527	100	81.5502	81.5502	81.5502	0.0232	232
15	K3280-04	BP-FW-MW03-20190608	84.7555	84.7555	100	84.794	84.794	84.794	0.0385	385
16	K3280-05	BP-HN-MW27I-20190608	63.8393	63.8393	100	63.8495	63.8495	63.8495	0.0102	102
17	K3280-06	BP-HN-MW29IR-20190608	104.6995	104.6995	100	104.7086	104.7086	104.7086	0.0091	91
18	K3280-07	BP-HN-MW29D-20190608	90.7979	90.7979	100	90.812	90.812	90.812	0.0141	141
19	K3280-08	BP-MH-SW4001-SOUTH-2019060	89.1666	89.1666	100	89.1924	89.1924	89.1924	0.0258	258
20	K3280-11	BP-TT-SW4002-20190608	74.4861	74.4861	100	78.5147	78.5147	78.5147	4.0286	40286
21	K3280-12	BP-TT-SW4003-20190608	82.47	82.47	100	82.4975	82.4975	82.4975	0.0275	275



TOTAL Dissolved Solids - SM2540C

Run Number: LB103323

SUPERVISOR: apatel

ANALYST: Aparna

Date: 06/12/2019

BalanceID: WC SC-3

Filter ID: 14466061

TEMP1 IN:	<u>103</u>	<u>06/10/2019</u>	<u>15:40</u>	TEMP OUT:	<u>105</u>	<u>06/12/2019</u>	<u>07:10</u>	OvenID1:	<u>WC Oven-2</u>	Thermo ID1:	<u>WET-OVEN#2</u>
TEMP2 IN:	<u>182</u>	<u>06/12/2019</u>	<u>07:15</u>	TEMP OUT:	<u>182</u>	<u>06/12/2019</u>	<u>09:00</u>	OvenID2:	<u>WC-OVEN-3</u>	Thermo ID2:	<u>WET-OVEN#3</u>
TEMP3 IN:	<u>182</u>	<u>06/12/2019</u>	<u>09:45</u>	TEMP OUT:	<u>182</u>	<u>06/12/2019</u>	<u>11:20</u>	OvenID3:	<u>WC-OVEN-3</u>	Thermo ID3:	<u>WET-OVEN#3</u>

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @180 (±2) °C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @180 (±2) °C (g)	Final Empty Dish+Sample weight after 1.5hr drying @180 (±2) °C (g)	Weight (g)	Result mg/L

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = 2nd Empty Dish + Sample weight after 1.5 hr drying @180 (±2) °C (g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)


WorkList Name : TDS061019

WorkList ID : 127176

Date : 06-10-2019 07:42:14

2B103323

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/14/2019	Water	K3279-02	TDS	Cool 4 deg C	TETR06	H23	BPS1-TT-MW312S-20190607	06/07/2019	SM2540 C
06/14/2019	Water	K3279-03	TDS	Cool 4 deg C	TETR06	H23	BPS1-TT-MW312L-20190607	06/07/2019	SM2540 C
06/14/2019	Water	K3279-04	TDS	Cool 4 deg C	TETR06	H23	BPS1-TT-MW302I2-20190607	06/07/2019	SM2540 C
06/14/2019	Water	K3279-05	TDS	Cool 4 deg C	TETR06	H23	BPS1-TT-MW307S-20190607	06/07/2019	SM2540 C
06/14/2019	Water	K3279-06	TDS	Cool 4 deg C	TETR06	H23	BPS1-TT-MW302D-20190607	06/07/2019	SM2540 C
06/14/2019	Water	K3279-07	TDS	Cool 4 deg C	TETR06	H23	BP-TT-MW203I-20190607	06/07/2019	SM2540 C
06/14/2019	Water	K3279-08	TDS	Cool 4 deg C	TETR06	H23	BPS1-TT-MW303I1-20190607	06/07/2019	SM2540 C
06/14/2019	Water	K3279-09	TDS	Cool 4 deg C	TETR06	H23	BP-EB03-20190607	06/07/2019	SM2540 C
06/15/2019	Water	K3280-02	TDS	Cool 4 deg C	TETR06	H41	BP-FW-MW01-20190608	06/08/2019	SM2540 C
06/15/2019	Water	K3280-03	TDS	Cool 4 deg C	TETR06	H41	BP-FW-MW02-20190608	06/08/2019	SM2540 C
06/15/2019	Water	K3280-04	TDS	Cool 4 deg C	TETR06	H41	BP-FW-MW03-20190608	06/08/2019	SM2540 C
06/15/2019	Water	K3280-05	TDS	Cool 4 deg C	TETR06	H41	BP-HN-MW27I-20190608	06/08/2019	SM2540 C
06/15/2019	Water	K3280-06	TDS	Cool 4 deg C	TETR06	H41	BP-HN-MW29IR-20190608	06/08/2019	SM2540 C
06/15/2019	Water	K3280-07	TDS	Cool 4 deg C	TETR06	H41	BP-HN-MW29D-20190608	06/08/2019	SM2540 C
06/15/2019	Water	K3280-08	TDS	Cool 4 deg C	TETR06	H41	BP-MH-SW4001-SOUTH-2019	06/08/2019	SM2540 C
06/15/2019	Water	K3280-11	TDS	Cool 4 deg C	TETR06	H41	BP-TT-SW4002-20190608	06/08/2019	SM2540 C
06/15/2019	Water	K3280-12	TDS	Cool 4 deg C	TETR06	H41	BP-TT-SW4003-20190608	06/08/2019	SM2540 C

Date/Time 06/12/19 10:30
 Received by: 
 Relinquished by: 

Date/Time 06/12/19 16:30
 Received by: 
 Relinquished by: 