



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB103329

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	06/11/2019	10:00
2	CAL2	1	Water	NA	NA	20.1	7.00	06/11/2019	10:01
3	CAL3	1	Water	NA	NA	20.1	10.04	06/11/2019	10:02
4	ICV	1	Water	NA	NA	20.1	6.99	06/11/2019	10:03
5	CCV1	1	Water	NA	NA	20.1	2.01	06/11/2019	10:05
6	K2640-02	1	Water	NA	NA	20.5	7.51	06/11/2019	10:06
7	K2640-02DUP	1	Water	NA	NA	20.6	7.52	06/11/2019	10:07
8	K2640-04	1	Water	NA	NA	21.2	4.91	06/11/2019	10:10
9	K2640-06	1	Water	NA	NA	20.8	6.76	06/11/2019	10:11
10	CCV2	1	Water	NA	NA	20.1	12.01	06/11/2019	10:15

LB103329

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K2640

WorkList ID : 127250

Date : 06-11-2019 09:46:26

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/04/2019	Water	K2640-02 ~1	pH	Cool 4 deg C	PSEG01	O31	30234	06/04/2019	9040C
06/04/2019	Water	K2640-04 ~1	pH	Cool 4 deg C	PSEG01	O31	30194	06/04/2019	9040C
06/04/2019	Water	K2640-06 ~1	pH	Cool 4 deg C	PSEG01	O31	21969	06/04/2019	9040C

Date/Time 06-11-19 9:50
Received by: JP
Relinquished by: PC

Date/Time 06-11-19 11:00
Received by: PC
Relinquished by: JP