

LB103337


ISM02.4

INSTRUMENT ID : CV1

Linear

 μ Abs.:

40235



A= 0.0000e+000

B= 2.4511e-004

C= -1.9026e-002

Rho= 0.9990633

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.00	0.000	-0.023	-0.023	-18	0.000	-18				
0.05	0.050									
0.20	0.200	0.186	-0.014	837	0.0 %	837				
2.50	2.500	2.605	0.105	10704	0.0 %	10704				
5.00	5.000	4.807	-0.193	19688	0.0 %	19688				
7.50	7.500	7.782	0.282	31827	0.0 %	31827				
10.0	10.000	9.844	-0.156	40239	0.0 %	40239				

6/10

-7
4
-4
4
-2

RS 06.11.19

LB103337

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type
	0 S0.0	-18	-		0 ISM02.4	PPB	6/11/2019 15:50	Std
	0.2 S0.2	837	-		0.2 ISM02.4	PPB	6/11/2019 15:53	Std
	2.5 S2.5	10704	-		2.5 ISM02.4	PPB	6/11/2019 15:56	Std
	5 S5.0	19688	-		5 ISM02.4	PPB	6/11/2019 15:58	Std
	7.5 S7.5	31827	-		7.5 ISM02.4	PPB	6/11/2019 16:00	Std
	10 S10.0	40239	-		10 ISM02.4	PPB	6/11/2019 16:12	Std
ICV	ICV	16702	4.0748	-	ISM02.4	PPB	6/11/2019 16:21	SMPL
ICB	ICB	-78	-0.0381	-	ISM02.4	PPB	6/11/2019 16:23	SMPL
CCV032	CCV032	19750	4.8219	-	ISM02.4	PPB	6/11/2019 16:26	SMPL
CCB032	CCB032	-101	-0.0438	-	ISM02.4	PPB	6/11/2019 16:28	SMPL
PB120521BL	PBW521	-26	-0.0254	-	ISM02.4	PPB	6/11/2019 16:30	SMPL
K3127-01	MH0AS2	229	0.0371	-	ISM02.4	PPB	6/11/2019 16:32	SMPL
K3127-02	MH0AS6	-13	-0.0222	-	ISM02.4	PPB	6/11/2019 16:35	SMPL
K3127-03	MH0AS9	29	-0.0119	-	ISM02.4	PPB	6/11/2019 16:37	SMPL
K3127-04	MH0AT4	4961	1.197	-	ISM02.4	PPB	6/11/2019 16:39	SMPL
K3127-05	MH0AU1	229	0.0371	-	ISM02.4	PPB	6/11/2019 16:42	SMPL
K3127-06	MH0AU5	485	0.0999	-	ISM02.4	PPB	6/11/2019 16:44	SMPL
K3127-07	MH0AU6	-58	-0.0332	-	ISM02.4	PPB	6/11/2019 16:46	SMPL
K3127-08	MH0AV3	-13	-0.0222	-	ISM02.4	PPB	6/11/2019 16:48	SMPL
K3127-09	MH0AV6	1273	0.293	-	ISM02.4	PPB	6/11/2019 16:51	SMPL
K3127-10	MH0AW5	-63	-0.0345	-	ISM02.4	PPB	6/11/2019 16:53	SMPL
K3127-11	MH0AW8	84	0.0016	-	ISM02.4	PPB	6/11/2019 16:55	SMPL
K3127-12	MH0AX4	-24	-0.0249	-	ISM02.4	PPB	6/11/2019 16:57	SMPL
K3127-13	MH0AX5	-53	-0.032	-	ISM02.4	PPB	6/11/2019 17:00	SMPL
K3127-14	MH0AX7	-53	-0.032	-	ISM02.4	PPB	6/11/2019 17:02	SMPL
K3127-15	MH0AZ0	1	-0.0188	-	ISM02.4	PPB	6/11/2019 17:04	SMPL
K3127-16	MH0AZ7	-40	-0.0288	-	ISM02.4	PPB	6/11/2019 17:06	SMPL
K3127-17	MH0AZ9	410	0.0815	-	ISM02.4	PPB	6/11/2019 17:09	SMPL
K3127-18	MH0BA2	-99	-0.0433	-	ISM02.4	PPB	6/11/2019 17:11	SMPL
K3127-19	MH0BA3	426	0.0854	-	ISM02.4	PPB	6/11/2019 17:13	SMPL
CCV033	CCV033	19744	4.8204	-	ISM02.4	PPB	6/11/2019 17:16	SMPL
CCB033	CCB033	-75	-0.0374	-	ISM02.4	PPB	6/11/2019 17:18	SMPL
K3127-20	MH0BB5	-17	-0.0232	-	ISM02.4	PPB	6/11/2019 17:20	SMPL
K3127-21	MH0BB5D	-32	-0.0269	-	ISM02.4	PPB	6/11/2019 17:22	SMPL
K3127-22	MH0BB5S	5006	1.208	-	ISM02.4	PPB	6/11/2019 17:25	SMPL
PB120521TB	LEB521	-95	-0.0423	-	ISM02.4	PPB	6/11/2019 17:27	SMPL
CCV034	CCV034	19597	4.7844	-	ISM02.4	PPB	6/11/2019 17:29	SMPL
CCB034	CCB034	-65	-0.035	-	ISM02.4	PPB	6/11/2019 17:31	SMPL