



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB103339

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/11/2019	15:00
2	CAL2	1	Water	NA	NA	20.2	7.01	06/11/2019	15:01
3	CAL3	1	Water	NA	NA	20.2	10.04	06/11/2019	15:02
4	ICV	1	Water	NA	NA	20.2	7.00	06/11/2019	15:05
5	CCV1	1	Water	NA	NA	20.1	2.01	06/11/2019	15:06
6	K3301-01	1	Water	NA	NA	22.6	5.29	06/11/2019	15:07
7	K3301-01DUP	1	Water	NA	NA	22.5	5.30	06/11/2019	15:08
8	CCV2	1	Water	NA	NA	20.2	12.01	06/11/2019	15:10

LP 103339

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K3301

WorkList ID : 127274

Date : 06-11-2019 14:47:47

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/06/2019	Water	K3301-01-17	pH	Cool 4 deg C	STER08	H11	POOL-WATER	06/06/2019	SM 4500-H B

Date/Time 06-11-19 14:50
 Received by: Jh
 Relinquished by: R

Date/Time 06-11-19 18:00
 Received by: R
 Relinquished by: Jh