



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB103340

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	06/11/2019	17:15
2	CAL2	1	Water	NA	NA	20.1	7.00	06/11/2019	17:16
3	CAL3	1	Water	NA	NA	20.1	10.02	06/11/2019	17:18
4	ICV	1	Water	NA	NA	20.2	7.01	06/11/2019	17:20
5	CCV1	1	Water	NA	NA	20.1	2.02	06/11/2019	17:21
6	K3159-02	1	Solid	20.02	20	22.6	8.24	06/11/2019	17:22
7	K3159-04	1	Solid	20.03	20	21.8	8.79	06/11/2019	17:25
8	K3265-09	1	Solid	20.02	20	21.6	9.14	06/11/2019	17:26
9	K3265-10	1	Solid	20.03	20	21.6	9.03	06/11/2019	17:28
10	K3265-11	1	Solid	20.05	20	21.1	10.93	06/11/2019	17:30
11	K3265-12	1	Solid	20.03	20	20.9	9.67	06/11/2019	17:32
12	K3282-02	1	Solid	20.01	20	21.5	7.40	06/11/2019	17:33
13	K3283-03	1	Solid	20.04	20	21.4	5.70	06/11/2019	18:00
14	K3283-03DUP	1	Solid	20.03	20	20.6	5.71	06/11/2019	18:01
15	CCV2	1	Water	NA	NA	20.2	12.02	06/11/2019	18:05

LB 103340

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity k3265

WorkList ID : 127227

Date : 06-11-2019 14:10:25

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/11/2019	Solid	K3159-02	Corrosivity	Cool 4 deg C	PSEG05	O31	OR-03-061119-A	06/11/2019	9045D
06/11/2019	Solid	K3159-04	Corrosivity	Cool 4 deg C	PSEG05	O31	OR-03-061119-B	06/11/2019	9045D
06/05/2019	Solid	K3265-09	Corrosivity	Cool 4 deg C	LIRO01	H23	B-01-(0-37)COMP	06/05/2019	9045D
06/05/2019	Solid	K3265-10	Corrosivity	Cool 4 deg C	LIRO01	H23	B-04-COMP	06/05/2019	9045D
06/06/2019	Solid	K3265-11	Corrosivity	Cool 4 deg C	LIRO01	H23	B-03-COMP	06/06/2019	9045D
06/07/2019	Solid	K3265-12	Corrosivity	Cool 4 deg C	LIRO01	H23	B-02-COMP	06/07/2019	9045D
06/10/2019	Solid	K3282-02	Corrosivity	Cool 4 deg C	PSEG01	O31	25082-2022	06/10/2019	9045D
06/10/2019	Solid	K3283-03	Corrosivity	Cool 4 deg C	PSEG01	K31	200030-200034	06/10/2019	9045D

Date/Time 06-11-19 16:00
 Received by: [Signature]
 Relinquished by: [Signature]

Date/Time 06-11-19 18:00
 Received by: [Signature]
 Relinquished by: [Signature]