



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB103341

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	06/11/2019	17:15
2	CAL2	1	Water	NA	NA	20.1	7.00	06/11/2019	17:16
3	CAL3	1	Water	NA	NA	20.1	10.02	06/11/2019	17:18
4	ICV	1	Water	NA	NA	20.2	7.01	06/11/2019	17:20
5	CCV1	1	Water	NA	NA	20.1	2.02	06/11/2019	18:10
6	K3303-05	1	Solid	20.03	20	20.6	7.94	06/11/2019	18:11
7	K3303-08	1	Solid	20.04	20	20.2	8.68	06/11/2019	18:12
8	K3303-08DUP	1	Solid	20.03	20	20.3	8.69	06/11/2019	18:14
9	CCV2	1	Water	NA	NA	20.2	12.02	06/11/2019	18:15

2B 103341

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH S K3308

WorkList ID : 127276

Date : 06-11-2019 16:51:47

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/10/2019	Solid	K3303-05-1A	pH	Cool 4 deg C	loui01	H51	BAT-B02-GEO	06/10/2019	9045D
06/11/2019	Solid	K3303-08-1A	pH	Cool 4 deg C	loui01	H51	BAT-B13-GEO	06/11/2019	9045D

Date/Time 06-11-19 18:00
 Received by: Jo
 Relinquished by: PC

Date/Time 06-11-19 18:00
 Received by: R
 Relinquished by: Jo